



Agilent 1200 Infinity Series Diode Array Detectors

User Manual



Agilent Technologies

Notices

© Agilent Technologies, Inc. 2010

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number

G4212-90010

Edition

06/2010

Printed in Germany

Agilent Technologies
Hewlett-Packard-Strasse 8
76337 Waldbronn

Warranty

The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed as “Commercial computer software” as defined in DFAR 252.227-7014 (June 1995), or as a “commercial item” as defined in FAR 2.101(a) or as “Restricted computer software” as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Agilent Technologies’ standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will

receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

For Research Use Only

In This Book

This manual covers
the Agilent 1290 Infinity Diode Array Detector (G4212A) and
the Agilent 1260 Infinity Diode Array Detector (G4212B).

Find information on other Agilent Diode Array Detectors in separate manuals.

1 Introduction

This chapter gives an introduction to the detector, instrument overview and internal connectors.

2 Site Requirements and Specifications

This chapter provides information on environmental requirements, physical and performance specifications.

3 Installing the Module

This chapter provides information on unpacking, checking on completeness, stack considerations and installation of the module.

4 LAN Configuration

This chapter provides information on connecting the module to the Agilent ChemStation PC.

5 Using the Module

This chapter provides information on how to set up the module for an analysis and explains the basic settings.

6 Optimizing the Detector

This chapter provides information on how to optimize the detector.

7 Troubleshooting and Diagnostics

Overview about the troubleshooting and diagnostic features.

8 Error Information

This chapter describes the meaning of error messages, and provides information on probable causes and suggested actions how to recover from error conditions.

9 Test Functions and Calibration

This chapter describes the tests for the module.

10 Maintenance

This chapter describes the maintenance of the module.

11 Parts and Materials for Maintenance

This chapter provides information on parts for maintenance.

12 Identifying Cables

This chapter provides information on cables used with the Agilent 1260 Infinity/1290 Infinity LC modules.

13 Appendix


This chapter provides addition information on safety, legal and web.

Contents

1	Introduction	9
	Overview of the Module	10
	Optical System	11
	Early Maintenance Feedback	17
	Instrument Layout	18
	Electrical Connections	19
	Interfaces	22
	Setting the 8-bit Configuration Switch	28
2	Site Requirements and Specifications	35
	Site Requirements	36
	Physical Specifications	40
	Performance Specifications	41
3	Installing the Module	45
	Unpacking the Module	46
	Optimizing the Stack Configuration	48
	Installing the Detector	57
	Flow Connections to the Detector	60
	Initial Recalibration	64
4	LAN Configuration	65
	What You Have to Do First	66
	TCP/IP parameter configuration	67
	Configuration Switch	68
	Initialization mode selection	69
	Link configuration selection	73
	Automatic Configuration with BootP	74
	Manual Configuration	84
	PC and Agilent ChemStation Setup	90

5	Using the Module	99
	Preparing the Detector	100
	Setting up the Detector with Agilent ChemStation	101
	Main Screens of the Detector with Agilent Instant Pilot (G4208A)	119
6	Optimizing the Detector	123
	Introduction	124
	Optimization Overview	125
	Optimizing for Sensitivity, Selectivity, Linearity and Dispersion	127
	Optimizing Selectivity	138
	Optimizing the Detector Regarding to the System	142
	Warm up of the Detector	149
7	Troubleshooting and Diagnostics	151
	Overview of the Module's Indicators and Test Functions	152
	Status Indicators	153
	Available Tests vs User Interfaces	155
	Agilent Lab Advisor Software	156
8	Error Information	157
	What Are Error Messages	158
	General Error Messages	159
	Detector Error Messages	166
9	Test Functions and Calibration	173
	Introduction	174
	Use of Max-Light Cartridge Test Cell	175
	Conditions of Detector	176
	Failing a Test	177
	Self-Test	178
	Intensity Test	180
	Cell Test	183
	Quick Noise Test	185
	ASTM Drift and Noise Test	188
	Slit Test (G4212A)	191
	Wavelength Verification Test	193

Wavelength Calibration	195
D/A Converter (DAC) Test	198
Dark Current Test	200
10 Maintenance	203
Warnings and Cautions	204
Introduction to Maintenance	206
Overview of Maintenance	207
Cleaning the Module	208
Replacing the Deuterium Lamp	209
Replacing the Max-Light Cartridge Cell	213
Cleaning of Max-Light Cartridge Cell	218
Storage of Max-Light Cartridge Cell	219
Drying the Leak Sensor	220
Replacing Leak Handling System Parts	221
Replacing Module Firmware	223
Information from Module's Assemblies	224
11 Parts and Materials for Maintenance	225
Overview of Maintenance Parts	226
Accessory Kits	228
12 Identifying Cables	229
Cable Overview	230
Analog Cables	232
Remote Cables	234
BCD Cables	237
CAN/LAN Cables	239
RS-232 Cable Kit	240
13 Appendix	241
Safety	242
The Waste Electrical and Electronic Equipment Directive	246
Radio Interference	247
Sound Emission	248
Flow Cell	249
Agilent Technologies on Internet	250



1

Introduction

Overview of the Module	10
Optical System	11
Lamp	12
Max-Light Cartridge Flow Cell	13
Slit Assembly	14
Grating and Diode Array	16
Early Maintenance Feedback	17
Instrument Layout	18
Electrical Connections	19
Rear view of the module	20
Information on Instrument Serial Number	21
Interfaces	22
Overview Interfaces	24
Setting the 8-bit Configuration Switch	28
Communication Settings for RS-232C	31
Special Settings	33

This chapter gives an introduction to the detector, instrument overview and internal connectors.



Overview of the Module

The detector is designed for highest optical performance, GLP compliance and easy maintenance. It includes the following features:

- Maximum of 160 Hz (G4212A) or 80 Hz (G4212B) data acquisition rate.
- Higher sensitivity for conventional LC as well as ultra fast applications by using next generation optical design.
- Increased sensitivity with 60 mm Max-Light cartridge flow cell.
- Optimized cell geometry for less peak dispersion for narrow bore applications.
- More reliable and robust peak integration process (automated) due to less baseline noise/drift/refractive index and thermal effects especially under ultra fast gradient conditions.
- RFID tracking technology is used for the UV-lamp and the Max-Light cartridge flow cells.
- Multiple wavelength and full spectral detection at 160 Hz (G4212A)/80 Hz (G4212B) sampling rate, keeping up with the analysis speed of ultra-fast LC.
- Programmable 1 – 8 nm slit (G4212A) or fixed 4 nm slit (G4212B) for rapid optimization of sensitivity, linearity and spectral resolution provides optimum incident light conditions .
- Improved Electronic temperature control (ETC) provides maximum baseline stability and practical sensitivity under fluctuating ambient temperature and humidity conditions.
- Additional diagnostic signals for temperature and lamp voltage monitoring.
- Easy exchange of flow cell by cartridge design.

For specifications, see “[Specifications](#)” on page 41.

Optical System

The optical system of the detector is shown in [Figure 1](#) on page 11.




Figure 1 Optical System of the Detector

1	UV-lamp
2	Lamp mirror
3	Flow cell
4	Fold mirror
5	Programmable (G4212A) or Fixed (G4212B) slit
6	Grating
7	Array

The illumination source is a deuterium-arc-discharge lamp [1] for the ultraviolet (UV) wavelength range. Its light is focused by a lamp mirror [2] onto the entrance of the Max-light cartridge flow cell [3] with optofluidic waveguides. The light leaves the Max-light cartridge flow cell at the other side and is focused by the fold mirror [4] through the slit assembly [5] onto a holographic grating [6] light being dispersed onto the diode array [7]. This allows simultaneous access to all wavelength information.

Lamp

The light source for the UV-wavelength range is a long-life UV-lamp with RFID tag. As a result of plasma discharge in low-pressure deuterium gas, the lamp emits light over the 190 nm to approximately 800 nm wavelength range.




Figure 2 UV-Lamp

Max-Light Cartridge Flow Cell

The detector allows easy access to flow cells via a cartridge. A variety of optional flow cells can be inserted using the same quick, simple mounting system. A Max-Light Cartridge Cell (10 mm, $V(\sigma)$ 1.0 μl) and a Max-Light Cartridge Cell (60 mm, $V(\sigma)$ 4 μL) are available. For testing of the detector, a Max-Light Cartridge Test Cell is available.

The optical principle of the Max-Light Cartridge cell is based on opto-fluidic waveguides. Nearly 100 % light transmission is achieved by utilizing total internal reflection in a non-coated silica fiber. Compromising refractive index and thermal effects are almost completely eliminated, resulting in significantly less baseline drift.




Figure 3 Max-Light Cartridge Flow Cell

Slit Assembly

Programmable Slit (G4212A)

The micro-slit system makes use of the mechanical properties of silicon combined with the precise structuring capabilities of bulk micro-machining. It combines the required optical functions — slit and shutter — in a simple and compact component. The slit width is directly controlled by the micro-processor of the instrument and can be set as method parameter.




Figure 4 Slit Assembly

The slit width influences the spectral resolution and noise.




Figure 5 Influence of slitwidth on resolution and noise level

Fixed Slit (G4212B)

The fixed slit combines the required optical functions - slit and shutter - in a simple and compact component. The slit width is directly controlled by the micro-processor of the instrument and is fixed to 4 nm.

Grating and Diode Array

The combination of dispersion and spectral imaging is accomplished by using a concave holographic grating. The grating separates the light beam into all its component wavelengths and reflects the light onto the photodiode array.

The diode array is a series of 1024 individual photodiodes and control circuits located on a ceramic carrier. It has a wavelength range from 190 – 640 nm and the sampling interval is ~ 0.5 nm.




Figure 6 Grating and diode array

Early Maintenance Feedback

Maintenance requires the exchange of components which are subject to wear or stress. Ideally, the frequency at which components are exchanged should be based on the intensity of usage of the module and the analytical conditions, and not on a predefined time interval. The early maintenance feedback (**EMF**) feature monitors the usage of specific components in the instrument, and provides feedback when the user-selectable limits have been exceeded. The visual feedback in the user interface provides an indication that maintenance procedures should be scheduled.

EMF Counters

EMF counters increment with use and can be assigned a maximum limit which provides visual feedback in the user interface when the limit is exceeded. Some counters can be reset to zero after the required maintenance procedure.

Using the EMF Counters

The user-settable **EMF** limits for the **EMF Counters** enable the early maintenance feedback to be adapted to specific user requirements. The useful maintenance cycle is dependent on the requirements for use. Therefore, the definition of the maximum limits need to be determined based on the specific operating conditions of the instrument.

Setting the EMF Limits

The setting of the **EMF** limits must be optimized over one or two maintenance cycles. Initially the default **EMF** limits should be set. When instrument performance indicates maintenance is necessary, take note of the values displayed by the **EMF counters**. Enter these values (or values slightly less than the displayed values) as **EMF** limits, and then reset the **EMF counters** to zero. The next time the **EMF counters** exceed the new **EMF** limits, the **EMF** flag will be displayed, providing a reminder that maintenance needs to be scheduled.

Instrument Layout

The industrial design of the module incorporates several innovative features. It uses Agilent's E-PAC concept for the packaging of electronics and mechanical assemblies. This concept is based upon the use of expanded polypropylene (EPP) layers of foam plastic spacers in which the mechanical and electronic boards components of the module are placed. This pack is then housed in a metal inner cabinet which is enclosed by a plastic external cabinet. The advantages of this packaging technology are:

- virtual elimination of fixing screws, bolts or ties, reducing the number of components and increasing the speed of assembly/disassembly,
- the plastic layers have air channels molded into them so that cooling air can be guided exactly to the required locations,
- the plastic layers help cushion the electronic and mechanical parts from physical shock, and
- the metal inner cabinet shields the internal electronics from electromagnetic interference and also helps to reduce or eliminate radio frequency emissions from the instrument itself.

Electrical Connections

- The CAN bus is a serial bus with high speed data transfer. The two connectors for the CAN bus are used for internal module data transfer and synchronization.
- One analog output provides signals for integrators or data handling systems.
- The REMOTE connector may be used in combination with other analytical instruments from Agilent Technologies if you want to use features such as start, stop, common shut down, prepare, and so on.
- With the appropriate software, the RS-232C connector may be used to control the module from a computer through a RS-232C connection. This connector is activated and can be configured with the configuration switch.
- The power input socket accepts a line voltage of 100 – 240 VAC \pm 10 % with a line frequency of 50 or 60 Hz. Maximum power consumption varies by module. There is no voltage selector on your module because the power supply has wide-ranging capability. There are no externally accessible fuses, because automatic electronic fuses are implemented in the power supply.

NOTE

Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.

Rear view of the module




Figure 7 Rear View of Detector – Electrical Connections and Label

NOTE

The CompactFlash Card Slot is not active yet. It may be used for future enhancements.

Information on Instrument Serial Number

Serial Number Information 1200 Series and 1290 Infinity

The serial number information on the instrument labels provide the following information:

CCYWWSSSSS	Format
CC	country of manufacturing <ul style="list-style-type: none"> • DE = Germany • JP = Japan • CN = China
YWW	year and week of last major manufacturing change, e.g. 820 could be week 20 of 1998 or 2008
SSSSS	real serial number

Serial Number Information 1260 Infinity

The serial number information on the instrument labels provide the following information:

CCXZZ00000	Format
CC	Country of manufacturing (DE Germany)
X	Alphabetic character A-Z (used by manufacturing)
ZZ	Alpha-numeric code 0-9, A-Z, where each combination unambiguously denotes a module (there can be more than one code for the same module)
00000	Serial number

Interfaces

The Agilent 1200 Infinity Series modules provide the following interfaces:

Table 1 Agilent 1200 Infinity Series Interfaces

Module	CAN	LAN/BCD (optional)	LAN (on-board)	RS-232	Analog	APG Remote	Special
Pumps							
G1310B Iso Pump G1311B Quat Pump G1311C Quat Pump VL G1312B Bin Pump G1312C Bin Pump VL 1376A Cap Pump G2226A Nano Pump	2	Yes	No	Yes	1	Yes	
G4220A/B Bin Pump	2	No	Yes	Yes	No	Yes	
G1361A Prep Pump	2	Yes	No	Yes	No	Yes	CAN-DC- OUT for CAN slaves
Samplers							
G1329B ALS G2260A Prep ALS	2	Yes	No	Yes	No	Yes	THERMOSTAT for G1330B
G1364B FC-PS G1364C FC-AS G1364D FC- μ S G1367E HiP ALS G1377A HiP micro ALS G2258A DL ALS	2	Yes	No	Yes	No	Yes	THERMOSTAT for G1330B CAN-DC- OUT for CAN slaves
G4226A ALS	2	Yes	No	Yes	No	Yes	
Detectors							
G1314B VWD VL G1314C VWD VL+	2	Yes	No	Yes	1	Yes	
G1314E/F VWD	2	No	Yes	Yes	1	Yes	

Table 1 Agilent 1200 Infinity Series Interfaces

Module	CAN	LAN/BCD (optional)	LAN (on-board)	RS-232	Analog	APG Remote	Special
G4212A/B DAD	2	No	Yes	Yes	1	Yes	
G1315C DAD VL+ G1365C MWD G1315D DAD VL G1365D MWD VL	2	No	Yes	Yes	2	Yes	
G1321B FLD G1362A RID	2	Yes	No	Yes	1	Yes	
G4280A ELSD	No	No	No	Yes	Yes	Yes	EXT Contact AUTOZERO
Others							
G1316A/C TCC	2	No	No	Yes	No	Yes	
G1322A DEG	No	No	No	No	No	Yes	AUX
G1379B DEG	No	No	No	Yes	No	No	AUX
G4227A Flex Cube	2	No	No	No	No	No	
G4240A CHIP CUBE	2	Yes	No	Yes	No	Yes	CAN-DC- OUT for CAN slaves THERMOSTAT for G1330A/B (NOT USED)

NOTE

The detector (DAD/MWD/FLD/VWD/RID) is the preferred access point for control via LAN. The inter-module communication is done via CAN.

- CAN connectors as interface to other modules
- LAN connector as interface to the control software
- RS-232C as interface to a computer
- REMOTE connector as interface to other Agilent products
- Analog output connector(s) for signal output

Overview Interfaces

CAN

The CAN is inter-module communication interface. It is a 2-wire serial bus system supporting high speed data communication and real-time requirement.

LAN

The modules have either an interface slot for an LAN card (e.g. Agilent G1369A/B LAN Interface) or they have an on-board LAN interface (e.g. detectors G1315C/D DAD and G1365C/D MWD). This interface allows the control of the module/system via a connected PC with the appropriate control software.

NOTE

If an Agilent detector (DAD/MWD/FLD/VWD/RID) is in the system, the LAN should be connected to the DAD/MWD/FLD/VWD/RID (due to higher data load). If no Agilent detector is part of the system, the LAN interface should be installed in the pump or autosampler.

RS-232C (Serial)

The RS-232C connector is used to control the module from a computer through RS-232C connection, using the appropriate software. This connector can be configured with the configuration switch module at the rear of the module. Refer to *Communication Settings for RS-232C*.

NOTE

There is no configuration possible on main boards with on-board LAN. These are pre-configured for

- 19200 baud,
- 8 data bit with no parity and
- one start bit and one stop bit are always used (not selectable).

The RS-232C is designed as DCE (data communication equipment) with a 9-pin male SUB-D type connector. The pins are defined as:

Table 2 RS-232C Connection Table

Pin	Direction	Function
1	In	DCD
2	In	RxD
3	Out	TxD
4	Out	DTR
5		Ground
6	In	DSR
7	Out	RTS
8	In	CTS
9	In	RI




Figure 8 RS-232 Cable

Analog Signal Output

The analog signal output can be distributed to a recording device. For details refer to the description of the module’s main board.

APG Remote

The APG Remote connector may be used in combination with other analytical instruments from Agilent Technologies if you want to use features as common shut down, prepare, and so on.

Remote control allows easy connection between single instruments or systems to ensure coordinated analysis with simple coupling requirements.

The subminiature D connector is used. The module provides one remote connector which is inputs/outputs (wired- or technique).

To provide maximum safety within a distributed analysis system, one line is dedicated to **SHUT DOWN** the system's critical parts in case any module detects a serious problem. To detect whether all participating modules are switched on or properly powered, one line is defined to summarize the **POWER ON** state of all connected modules. Control of analysis is maintained by signal readiness **READY** for next analysis, followed by **START** of run and optional **STOP** of run triggered on the respective lines. In addition **PREPARE** and **START REQUEST** may be issued. The signal levels are defined as:

- standard TTL levels (0 V is logic true, + 5.0 V is false),
- fan-out is 10,
- input load is 2.2 kOhm against + 5.0 V, and
- output are open collector type, inputs/outputs (wired- or technique).

NOTE

All common TTL circuits operate with a 5 V power supply. A TTL signal is defined as "low" or L when between 0 V and 0.8 V and "high" or H when between 2.0 V and 5.0 V (with respect to the ground terminal).

Table 3 Remote Signal Distribution

Pin	Signal	Description
1	DGND	Digital ground
2	PREPARE	(L) Request to prepare for analysis (for example, calibration, detector lamp on). Receiver is any module performing pre-analysis activities.
3	START	(L) Request to start run / timetable. Receiver is any module performing run-time controlled activities.
4	SHUT DOWN	(L) System has serious problem (for example, leak: stops pump). Receiver is any module capable to reduce safety risk.
5		Not used
6	POWER ON	(H) All modules connected to system are switched on. Receiver is any module relying on operation of others.
7	READY	(H) System is ready for next analysis. Receiver is any sequence controller.
8	STOP	(L) Request to reach system ready state as soon as possible (for example, stop run, abort or finish and stop injection). Receiver is any module performing run-time controlled activities.
9	START REQUEST	(L) Request to start injection cycle (for example, by start key on any module). Receiver is the autosampler.

Special Interfaces

Some modules have module specific interfaces/connectors. They are described in the module documentation.

Setting the 8-bit Configuration Switch

Setting the 8-bit Configuration Switch (On-Board LAN)

The 8-bit configuration switch is located at the rear of the module. Switch settings provide configuration parameters for LAN, serial communication protocol and instrument specific initialization procedures.

All modules with on-board LAN, e.g. G1315/65C/D, G1314D/E/F, G4212A/B, G4220A:

- Default is ALL switches DOWN (best settings) - Bootp mode for LAN.
- For specific LAN modes switches 3-8 must be set as required.
- For boot/test modes switches 1+2 must be UP plus required mode.




Figure 9 Location of Configuration Switch (example shows a G4212A DAD)

NOTE

To perform any LAN configuration, SW1 and SW2 must be set to OFF. For details on the LAN settings/configuration refer to chapter LAN Configuration.

Table 4 8-bit Configuration Switch (with on-board LAN)

	Mode		Function					
	SW 1	SW 2	SW 3	SW 4	SW 5	SW 6	SW 7	SW 8
LAN	0	0	Link Configuration			Init Mode Selection		
Auto-negotiation			0	x	x	x	x	x
10 MBit, half-duplex			1	0	0	x	x	x
10 MBit, full-duplex			1	0	1	x	x	x
100 MBit, half-duplex			1	1	0	x	x	x
100 MBit, full-duplex			1	1	1	x	x	x
Bootp			x	x	x	0	0	0
Bootp & Store			x	x	x	0	0	1
Using Stored			x	x	x	0	1	0
Using Default			x	x	x	0	1	1
TEST	1	1	System					NVRAM
Boot Resident System			1					x
Revert to Default Data (Coldstart)			x	x	x			1

Legend:

0 (switch down), 1 (switch up), x (any position)

NOTE

When selecting the mode TEST, the LAN settings are: Auto-Negotiation & Using Stored.

NOTE

For explanation of "Boot Resident System" and "Revert to Default Data (Coldstart)" refer to ["Special Settings"](#) on page 33.

Setting the 8-bit Configuration Switch (without On-Board LAN)

The 8-bit configuration switch is located at the rear of the module.

Modules that do not have their own LAN interface (e.g. the TCC) can be controlled through the LAN interface of another module and a CAN connection to that module.




Figure 10 Configuration switch (settings depend on configured mode)

All modules without on-board LAN:

- default is ALL DIPS DOWN (best settings) - Bootp mode for LAN
- for boot/test modes DIPS 1+2 must be UP plus required mode

Switch settings provide configuration parameters for GPIB address, serial communication protocol and instrument specific initialization procedures.

NOTE

With the introduction of the Agilent 1260 Infinity, all GPIB interfaces have been removed. The preferred communication is LAN.

NOTE

The following tables represent the configuration switch settings for the modules without on-board LAN only.

Table 5 8-bit Configuration Switch (without on-board LAN)

Mode Select	1	2	3	4	5	6	7	8
RS-232C	0	1	Baudrate			Data Bits	Parity	
Reserved	1	0	Reserved					
TEST/BOOT	1	1	RSVD	SYS		RSVD	RSVD	FC

NOTE

The LAN settings are done on the LAN Interface Card G1369A/B. Refer to the documentation provided with the card.

Communication Settings for RS-232C

The communication protocol used in the column compartment supports only hardware handshake (CTS/RTR).

Switches 1 in down and 2 in up position define that the RS-232C parameters will be changed. Once the change has been completed, the column instrument must be powered up again in order to store the values in the non-volatile memory.

Table 6 Communication Settings for RS-232C Communication (without on-board LAN)

Mode Select	1	2	3	4	5	6	7	8
RS-232C	0	1	Baudrate			Data Bits	Parity	

Use the following tables for selecting the setting which you want to use for RS-232C communication. The number 0 means that the switch is down and 1 means that the switch is up.

Table 7 Baudrate Settings (without on-board LAN)

Switches			Baud Rate	Switches			Baud Rate
3	4	5		3	4	5	
0	0	0	9600	1	0	0	9600
0	0	1	1200	1	0	1	14400
0	1	0	2400	1	1	0	19200
0	1	1	4800	1	1	1	38400

Table 8 Data Bit Settings (without on-board LAN)

Switch 6	Data Word Size
0	7 Bit Communication
1	8 Bit Communication

Table 9 Parity Settings (without on-board LAN)

Switches		Parity
7	8	
0	0	No Parity
1	0	Odd Parity
1	1	Even Parity

One start bit and one stop bit are always used (not selectable).

Per default, the module will turn into 19200 baud, 8 data bit with no parity.

Special Settings

The special settings are required for specific actions (normally in a service case).

NOTE

The tables include both settings for modules – with on-board LAN and without on-board LAN. They are identified as LAN and no LAN.

Boot-Resident

Firmware update procedures may require this mode in case of firmware loading errors (main firmware part).

If you use the following switch settings and power the instrument up again, the instrument firmware stays in the resident mode. It is not operable as a module. It only uses basic functions of the operating system for example, for communication. In this mode the main firmware can be loaded (using update utilities).

Table 10 Boot Resident Settings (without on-board LAN)

	Mode Select	SW1	SW2	SW3	SW4	SW5	SW6	SW7	SW8
LAN	TEST/BOOT	1	1	1	0	0	0	0	0
No LAN	TEST/BOOT	1	1	0	0	1	0	0	0

Forced Cold Start

A forced cold start can be used to bring the module into a defined mode with default parameter settings.

CAUTION

Loss of data

Forced cold start erases all methods and data stored in the non-volatile memory. Exceptions are diagnosis and repair log books which will not be erased.

→ Save your methods and data before executing a forced cold start.

If you use the following switch settings and power the instrument up again, a forced cold start has been completed.

Table 11 Forced Cold Start Settings (without on-board LAN)

	Mode Select	SW1	SW2	SW3	SW4	SW5	SW6	SW7	SW8
LAN	TEST/BOOT	1	1	0	0	0	0	0	1
No LAN	TEST/BOOT	1	1	0	0	1	0	0	1



2

Site Requirements and Specifications

Site Requirements	36
Physical Specifications	40
Performance Specifications	41
Specifications	41
Specification Conditions	43

This chapter provides information on environmental requirements, physical and performance specifications.



Site Requirements

A suitable environment is important to ensure optimal performance of the module.

Power Consideration

The module power supply has wide ranging capabilities and accepts any line voltage in the range mentioned in [Table 12](#) on page 40. Consequently, there is no voltage selector in the rear of the module. There are also no externally accessible fuses, because automatic electronic fuses are implemented in the power supply.

WARNING

Module is partially energized when switched off, as long as the power cord is plugged in.

Repair work at the module can lead to personal injuries, e.g. shock hazard, when the cover is opened and the module is connected to power.

- Make sure that it is always possible to access the power plug.
- Remove the power cable from the instrument before opening the cover.
- Do not connect the power cable to the Instrument while the covers are removed.

WARNING

Incorrect line voltage at the module

Shock hazard or damage of your instrument can result if the devices are connected to line voltage higher than specified.

- Connect your module to the specified line voltage.
-

CAUTION

Unaccessible power plug.

In case of emergency it must be possible to disconnect the instrument from the power line at any time.

- Make sure the power connector of the instrument can be easily reached and unplugged.
 - Provide sufficient space behind the power socket of the instrument to unplug the cable.
-

Power Cords

Different power cords are offered as options with the module. The female end of all power cords is identical. It plugs into the power-input socket at the rear. The male end of each power cord is different and designed to match the wall socket of a particular country or region.

WARNING

Absence of ground connection or use of unspecified power cord

The absence of ground connection or the use of unspecified power cord can lead to electric shock or short circuit.

- Never operate your instrumentation from a power outlet that has no ground connection.
 - Never use a power cord other than the Agilent Technologies power cord designed for your region.
-

WARNING

Use of unsupplied cables

Using cables not supplied by Agilent Technologies can lead to damage of the electronic components or personal injury.

- Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.
-

WARNING

Unintended use of supplied power cords

Using power cords for unintended purposes can lead to personal injury or damage of electronic equipment.

- Never use the power cords that Agilent Technologies supplies with this instrument for any other equipment.
-

Bench Space

The module dimensions and weight (see [Table 12](#) on page 40) allow you to place the module on almost any desk or laboratory bench. It needs an additional 2.5 cm (1.0 inches) of space on either side and approximately 8 cm (3.1 inches) in the rear for air circulation and electric connections.

If the bench should carry an Agilent system, make sure that the bench is designed to bear the weight of all modules.

The module should be operated in a horizontal position.

Environment

Your module will work within the specifications at ambient temperatures and relative humidity described in [Table 12](#) on page 40.

ASTM drift tests require a temperature change below 2 °C/hour (3.6 F/hour) over one hour period. Our published drift specification (refer also to “[Specifications](#)” on page 41) is based on these conditions. Larger ambient temperature changes will result in larger drift.

Better drift performance depends on better control of the temperature fluctuations. To realize the highest performance, minimize the frequency and the amplitude of the temperature changes to below 1 °C/hour (1.8 F/hour). Turbulences around one minute or less can be ignored.

CAUTION

Condensation within the module

Condensation will damage the system electronics.

- Do not store, ship or use your module under conditions where temperature fluctuations could cause condensation within the module.
- If your module was shipped in cold weather, leave it in its box and allow it to warm slowly to room temperature to avoid condensation.

NOTE

This module is designed to operate in a typical electromagnetic environment, i.e. where RF transmitters such as mobile telephones may not be used in close proximity.

Physical Specifications

Table 12 Physical Specifications

Type	Specification	Comments
Weight	11.5 kg (26 lbs)	
Dimensions (height × width × depth)	140 x 345 x 435 mm (5.5 x 13.5 x 17 inches)	
Line voltage	100 – 240 VAC, ± 10%	Wide-ranging capability
Line frequency	50 or 60 Hz, ± 5%	
Power consumption	160 VA / 130 W / 444 BTU	Maximum
Ambient operating temperature	4–40 °C (39–104 °F)	
Ambient non-operating temperature	-40–70 °C (-4–158 °F)	
Humidity	Operating: < 80 % Non-Operating: < 95 %	Non-condensing
Operating Altitude	Up to 2000 m (6562 ft)	
Non-operating altitude	Up to 4600 m (15091 ft)	For storing the module
Safety standards: IEC, CSA, UL	Installation Category II, Pollution Degree 2	For indoor use only.

Performance Specifications

Specifications

Table 13 Performance Specifications G4212A/B

Type	Specification	Comments
Detection type	1024-element photodiode array	
Light source	Deuterium lamp	Equipped with RFID tag that holds lamp typical information.
Wavelength range	190 – 640 nm	
Short term noise (ASTM) Single and Multi-Wavelength	$< \pm 3 \times 10^{-6}$ AU at 230 nm/4 nm	see "Specification Conditions" below
Drift	$< 0.5 \times 10^{-3}$ AU/hr at 230 nm	see "Specification Conditions" below
Linear absorbance range	> 2.0 AU (5 %) at 265 nm	see "Specification Conditions" below
Wavelength accuracy	± 1 nm	After recalibration with deuterium lines
Wavelength bunching	2 – 400 nm	Programmable in steps of 1 nm
Slit width	G4212A: 1, 2, 4, 8 nm G4212B: 4 nm	Programmable slit Fixed slit
Diode width	~ 0.5 nm	
Signal data rate	up to 160 Hz (G4212A) or 80 Hz (G4212B)	
Spectra Data rate	up to 160 Hz (G4212A) or 80 Hz (G4212B)	
Flow cells	Max-Light Cartridge Cell (10 mm, V(σ) 1.0 μ l), 60 bar (870 psi) pressure maximum Max-Light Cartridge Cell (60 mm), V(σ) 4.0 μ L), 60 bar (870 psi) pressure maximum Max-Light Cartridge Test Cell	pH range 1.0 —12.5 (solvent dependent) Cartridge type, equipped with RFID tags that holds cell typical information.

2 Site Requirements and Specifications

Performance Specifications

Table 13 Performance Specifications G4212A/B

Type	Specification	Comments
Control and data evaluation	Data System 1 Agilent ChemStation for LC 2 EZChrom Elite 3 MassHunter	For G4212A: 1 B.04.02 or above 2 3.3.2 SP1 or above 3 B.02.01 SP1 or above For G4212B: 1 B.04.02 DSP3 or above 2 3.3.2 SP2 or above 3 B.04.00 and B.03.01 SP2 or above
Local Control	Agilent Instant Pilot (G4208A)	B.02.11 or above
Test and diagnostic software	Agilent LabAdvisor	B.01.03 SP4 or above
Analog outputs	Recorder/integrator: 100 mV or 1 V, output range 0.001 – 2 AU, one output	
Communications	Controller-area network (CAN), RS-232C, APG Remote: ready, start, stop and shut-down signals, LAN	
Safety and maintenance	Extensive diagnostics, error detection and display (through control module and ChemStation), leak detection, safe leak handling, leak output signal for shutdown of pumping system. Low voltages in major maintenance areas.	
GLP features	Early maintenance feedback (EMF) for continuous tracking of instrument usage in terms of lamp burn time with user-setable limits and feedback messages. Electronic records of maintenance and errors. Verification of wavelength accuracy with the emission lines of the deuterium lamp.	
Housing	All materials recyclable.	

Specification Conditions

ASTM: “Standard Practice for Variable Wavelength Photometric Detectors Used in Liquid Chromatography”.

Reference conditions:

- Wavelength: 230 nm/4 nm with Reference Wavelength 360 nm/100 nm, Slitwidth 4 nm, TC 2 s, (or with $RT = 2.2 * TC$), ASTM
- Max-Light Cartridge Cell (10 mm, $V(\sigma)$ 1.0 μ l) with flow of 0.5 ml/min LC grade water or Max-Light Cartridge Test Cell

Linearity:

Linearity is measured with caffeine at 265 nm/4 nm with slit width 4 nm and TC 1 s (or with RT 2 s) with Max-Light Cartridge Cell (10 mm, $V(\sigma)$ 1 μ l) > 2.0 AU (5 %) [typical 2.5 AU (5 %)] .

NOTE

The specifications are based on the standard RFID tag lamp (5190-0917) and may be not achieved when other lamp types or aged lamps are used.

ASTM drift tests require a temperature change below 2 °C/hour (3.6 F/hour) over one hour period. Our published drift specification is based on these conditions. Larger ambient temperature changes will result in larger drift.

Better drift performance depends on better control of the temperature fluctuations. To realize the highest performance, minimize the frequency and the amplitude of the temperature changes to below 1 °C/hour (1.8 F/hour). Turbulences around one minute or less can be ignored.


Performance tests should be done with a completely warmed up optical unit (> two hours). ASTM measurements require that the detector should be turned on at least 24 h before start of testing.

Time Constant versus Response Time

According to ASTM E1657-98 „Standard Practice of Testing Variable-Wavelength Photometric Detectors Used in Liquid Chromatography” the time constant is converted to response time by multiplying by the factor 2.2.

2 Site Requirements and Specifications

Performance Specifications



3 Installing the Module

Unpacking the Module	46
Damage to the module	46
Delivery Checklist	47
Detector Accessory Kit Contents	47
Optimizing the Stack Configuration	48
One Stack Configuration	49
Two Stack Configuration	53
Installing the Detector	57
Flow Connections to the Detector	60
Initial Recalibration	64

This chapter provides information on unpacking, checking on completeness, stack considerations and installation of the module.



Unpacking the Module

Damage to the module

Damaged Packaging

If the delivery packaging shows signs of external damage, please call your Agilent Technologies sales and service office immediately. Inform your service representative that the instrument may have been damaged during shipment.

CAUTION

"Defective on arrival" problems

If there are signs of damage, please do not attempt to install the module. Inspection by Agilent is required to evaluate if the instrument is in good condition or damaged.

- Notify your Agilent sales and service office about the damage.
 - An Agilent service representative will inspect the instrument at your site and initiate appropriate actions.
-

Condensation

CAUTION

Condensation within the module

Condensation will damage the system electronics.

- Do not store, ship or use your module under conditions where temperature fluctuations could cause condensation within the module.
 - If your module was shipped in cold weather, leave it in its box and allow it to warm slowly to room temperature to avoid condensation.
-

Delivery Checklist

Ensure all parts and materials have been delivered with the module. The delivery checklist is shown below. Please report missing or damaged parts to your local Agilent Technologies sales and service office.

Table 14 Detector Checklist

Description	Quantity
Detector	1
Power cable	1
Cross-over network cable	1
Twisted pair network cable	1
Max-Light Cartridge Cell (10 mm, V(σ) 1.0 μ L) or (60 mm, V(σ) 4.0 μ L)	1
User Manual	on Documentation CD (part of the shipment - not module specific)
Accessory kit	1

Detector Accessory Kit Contents

Detector Accessory Kit Contents (p/n G4212-68755)

p/n	Description
5062-2462	Teflon Tubing flexible i.d. 0.8 mm, o.d. 1.6 mm, 2 m, re-order 5 m (flow cell to waste)
5063-6527	Tubing assembly, i.d. 6 mm, o.d. 9 mm, 1.2 m (to waste)
5042-9954	Tubing clip (2x), re-order 4/pk
0100-1516	Fitting male PEEK, 2/pk
5067-4660	Inlet Capillary SST 0.12 mm I.D., 220 mm long
5181-1516	CAN cable, Agilent module to module, 0.5 m

Optimizing the Stack Configuration

If your module is part of a complete Agilent 1260 Infinity/1290 Infinity LC System, you can ensure optimum performance by installing the following configurations. These configurations optimize the system flow path, ensuring minimum delay volume.

For other possible configurations, please refer to the Agilent 1260 Infinity/1290 Infinity LC System Manual.

One Stack Configuration

One Stack Configuration for Agilent 1260 Infinity LC

Ensure optimum performance by installing the modules of the Agilent 1260 Infinity LC System in the following configuration (See [Figure 11](#) on page 49 and [Figure 12](#) on page 50). This configuration optimizes the flow path for minimum delay volume and minimizes the bench space required.




Figure 11 Recommended Stack Configuration for 1260 (Front View)

3 Installing the Module

Optimizing the Stack Configuration




Figure 12 Recommended Stack Configuration for 1260 (Rear View)

One Stack Configuration for Agilent 1290 Infinity LC

Ensure optimum performance by installing the modules of the Agilent 1290 Infinity LC System in the following configuration (See [Figure 13](#) on page 51 and [Figure 14](#) on page 52). This configuration optimizes the flow path for minimum delay volume and minimizes the bench space required.

The Agilent 1290 Infinity Binary Pump should always be installed at the bottom of the stack.




Figure 13 Recommended Stack Configuration for 1290 (Front View)

3 Installing the Module

Optimizing the Stack Configuration




Figure 14 Recommended Stack Configuration (Rear View)

Two Stack Configuration

Two Stack Configuration for Agilent 1260 Infinity LC

To avoid excessive height of the stack when the autosampler thermostat is added to the system it is recommended to form two stacks. Some users prefer the lower height of this arrangement even without the autosampler thermostat. A slightly longer capillary is required between the pump and autosampler. (See [Figure 15](#) on page 53 and [Figure 16](#) on page 54).




Figure 15 Recommended Two Stack Configuration for 1260 (Front View)

3 Installing the Module

Optimizing the Stack Configuration




Figure 16 Recommended Two Stack Configuration for 1260 (Rear View)

Two Stack Configuration for Agilent 1290 Infinity LC

To avoid excessive height of the stack when the autosampler thermostat is added to the system it is recommended to form two stacks. Some users prefer the lower height of this arrangement even without the autosampler thermostat. A slightly longer capillary is required between the pump and autosampler. (See [Figure 17](#) on page 55 and [Figure 18](#) on page 56).




Figure 17 Recommended Two Stack Configuration for 1290 (Front View)

3 Installing the Module

Optimizing the Stack Configuration




Figure 18 Recommended Two Stack Configuration for 1290 (Rear View)

Installing the Detector

Parts required

#	Description
1	Detector
1	Power cord
1	LAN cable (cross-over or twisted pair network cable)

Other cables see below and section “[Cable Overview](#)” on page 230.

Software required

Instant Pilot and/or ChemStation with the appropriate revisions, see [Table 13](#) on page 41.

Preparations

Locate bench space
Provide power connections
Unpack the module

WARNING

Module is partially energized when switched off, as long as the power cord is plugged in.

Repair work at the module can lead to personal injuries, e.g. shock hazard, when the cover is opened and the module is connected to power.

- Make sure that it is always possible to access the power plug.
- Remove the power cable from the instrument before opening the cover.
- Do not connect the power cable to the Instrument while the covers are removed.

3 Installing the Module

Installing the Detector

- 1 Note the MAC address of the LAN interface (rear of the module, under the configuration switch, see [Figure 21](#) on page 66). It's required for LAN Configuration (see Chapter "[LAN Configuration](#)").




Figure 19 Rear View of Detector – Electrical Connections and Label

- 2 Set the Configuration Switch according the required initialization mode (default, fixed IP or Bootp), see Chapter "[LAN Configuration](#)".
- 3 Place the module in the stack, see "[Optimizing the Stack Configuration](#)" on page 48 .
- 4 Ensure the line power switch at the front of the module is OFF.
- 5 Connect the power cable to the power connector at the rear of the module.
- 6 Connect the CAN cable to other Agilent modules.
- 7 Connect the LAN cable (e.g. from an Agilent ChemStation as controller) to the detector's LAN connector.
- 8 Connect the analog cable (optional).
- 9 Connect the APG remote cable (optional) for non-Agilent instruments.
- 10 Turn on power by pushing the button at the lower left hand side of the module. The status LED should be green.

NOTE

The module is turned on when the line power switch is pressed and the green indicator lamp is illuminated. The module is turned off when the line power switch is protruding and the green light is off.

NOTE

The module was shipped with default configuration settings. To change these settings see chapter "*LAN Configuration*".

NOTE

After turn-on of the detector, it goes through a cycle of different states of heating up the optical unit and controlling the temperature. This is described in "[Warm up of the Detector](#)" on page 149.

Give the optical unit enough time to warm-up and stabilize (> 60 minutes).

3 Installing the Module

Flow Connections to the Detector


Flow Connections to the Detector

Parts required	#	Description
	1	System
	1	Max-Light cartridge flow cell
	1	Capillaries and tubing from <i>Accessory Kit</i> .


NOTE

This procedure shows the detector outside of a system. In an Agilent 1260 Infinity Liquid Chromatograph, the detector is located below a G1316 TCC on the bench, see [“Optimizing the Stack Configuration”](#) on page 48.


1 Remove the front cover.




2 Remove the black hoods from the cell interfaces (in/out) and store them in the plastic case provided with the Max-Light Cartridge Flow Cell.




3 Locate the flow cell cartridge.




4 Unlock the flow cell cartridge by pulling the lever to the front.



5 The lever should be in the final down position.




6 Pull the flow cell cartridge completely out towards the front.




3 Installing the Module

Flow Connections to the Detector


- 7** Remove the black hoods from the cell interfaces (in/out) and insert the cell into the cell cartridge holder.




- 8** Slide the cell cartridge holder completely into the module.




- 9** Lift the two levers into the upper final position to fix the cell.




- 10** Remove the plugs from the CELL-IN and CELL-OUT (keep them safe).



11 Connect the inlet capillary to CELL-IN (left) and the waste tubing to CELL-OUT (right).



12 Close the front cover.



13 Route the inlet capillary and waste tubing towards their destinations.

NOTE

The detector should be operated with the front cover in place to protect the flow cell area against strong drafts from the outside.


NOTE

If the flow cell is replaced by a different flow cell, it should be flushed with isopropanol and the CELL-IN and CELL-OUT should be closed with the plugs.

Initial Recalibration

The detector has been calibrated with a factory flow cell initially. After installation of the detector with the delivered or a new Max-Light Cartridge flow cell and an initial warm-up of at least 2 hours, a recalibration should be performed ([“Wavelength Calibration”](#) on page 195). This recalibration will correct slight changes due to

- significant environmental condition changes (temperature, humidity) during transport and storage,
- significant environmental condition changes (temperature, humidity) in final location and
- variances between factory test cell and the installed flow cell.



4 LAN Configuration

What You Have to Do First	66
TCP/IP parameter configuration	67
Configuration Switch	68
Initialization mode selection	69
Link configuration selection	73
Automatic Configuration with BootP	74
About Agilent BootP Service	74
How BootP Service Works	75
Situation: Cannot Establish LAN Communication	75
Installation of BootP Service	76
Two Methods to Determine the MAC Address	78
Assigning IP Addresses Using the Agilent BootP Service	79
Changing the IP Address of an Instrument Using the Agilent BootP Service	82
Manual Configuration	84
With Telnet	85
With the Instant Pilot (G4208A)	89
PC and Agilent ChemStation Setup	90
PC Setup for Local Configuration	90
Agilent ChemStation Setup	93

This chapter provides information on connecting the module to the Agilent ChemStation PC.



What You Have to Do First

The module has an on-board LAN communication interface.

- 1 Note the MAC (Media Access Control) address for further reference. The MAC or hardware address of the LAN interfaces is a world wide unique identifier. No other network device will have the same hardware address. The MAC address can be found on a label at the rear of the module underneath the configuration switch (see [Figure 21](#) on page 66).




Figure 20 MAC-Label

- 2 Connect the instrument's LAN interface (see [Figure 21](#) on page 66) to
 - the PC network card using a crossover network cable (point-to-point) or
 - a hub or switch using a standard LAN cable.




Figure 21 Location of LAN interfaces and MAC label

TCP/IP parameter configuration

To operate properly in a network environment, the LAN interface must be configured with valid TCP/IP network parameters. These parameters are:

- IP address
- Subnet Mask
- Default Gateway

The TCP/IP parameters can be configured by the following methods:

- by automatically requesting the parameters from a network-based BOOTP Server (using the so-called Bootstrap Protocol)
- by manually setting the parameters using Telnet
- by manually setting the parameters using the Instant Pilot (G4208A)

The LAN interface differentiates between several initialization modes. The initialization mode (short form 'init mode') defines how to determine the active TCP/IP parameters after power-on. The parameters may be derived from a Bootp cycle, non-volatile memory or initialized with known default values. The initialization mode is selected by the configuration switch, see [Table 18](#) on page 73.

Configuration Switch

The configuration switch can be accessed at the rear of the module.




Figure 22 Location of Configuration Switch

The module is shipped with all switches set to OFF, as shown above.

NOTE

To perform any LAN configuration, SW1 and SW2 must be set to OFF.


Table 15 Factory Default Settings

Initialization ('Init') Mode	Bootp, all switches down. For details see "Initialization mode selection" on page 69
Link Configuration	speed and duplex mode determined by auto-negotiation, for details see "Link configuration selection" on page 73

Initialization mode selection

The following initialization (init) modes are selectable:

Table 16 Initialization Mode Switches

	SW 6	SW 7	SW 8	Init Mode
	OFF	OFF	OFF	Bootp
	OFF	OFF	ON	Bootp & Store
	OFF	ON	OFF	Using Stored
	OFF	ON	ON	Using Default

Bootp

When the initialization mode **Bootp** is selected, the module tries to download the parameters from a **Bootp** Server. The parameters obtained become the active parameters immediately. They are not stored to the non-volatile memory of the module. Therefore, the parameters are lost with the next power cycle of the module.




Figure 23 Bootp (Principle)

Bootp & Store

When **Bootp & Store** is selected, the parameters obtained from a **Bootp** Server become the active parameters immediately. In addition, they are stored to the non-volatile memory of the module. Thus, after a power cycle they are still available. This enables a kind of bootp once configuration of the module.

Example: The user may not want to have a **Bootp** Server be active in his network all the time. But on the other side, he may not have any other configuration method than **Bootp**. In this case he starts the **Bootp** Server temporarily, powers on the module using the initialization mode **Bootp & Store**, waits for the **Bootp** cycle to be completed, closes the **Bootp** Server and powers off the module. Then he selects the initialization mode Using Stored and powers on the module again. From now on, he is able to establish the TCP/IP connection to the module with the parameters obtained in that single **Bootp** cycle.




Figure 24 Bootp & Store (Principle)

NOTE

Use the initialization mode **Bootp & Store** carefully, because writing to the non-volatile memory takes time. Therefore, when the module shall obtain its parameters from a **Bootp** Server every time it is powered on, the recommended initialization mode is **Bootp**!

Using Stored

When initialization mode **Using Stored** is selected, the parameters are taken from the non-volatile memory of the module. The TCP/IP connection will be established using these parameters. The parameters were configured previously by one of the described methods.




Figure 25 Using Stored (Principle)

Using Default

When **Using Default** is selected, the factory default parameters are taken instead. These parameters enable a TCP/IP connection to the LAN interface without further configuration, see [Table 17](#) on page 72.




Figure 26 Using Default (Principle)

NOTE

Using the default address in your local area network may result in network problems. Take care and change it to a valid address immediately.

Table 17 Using Default Parameters

IP address:	192.168.254.11
Subnet Mask:	255.255.255.0
Default Gateway	not specified

Since the default IP address is a so-called local address, it will not be routed by any network device. Thus, the PC and the module must reside in the same subnet.

The user may open a Telnet session using the default IP address and change the parameters stored in the non-volatile memory of the module. He may then close the session, select the initialization mode Using Stored, power-on again and establish the TCP/IP connection using the new parameters.

When the module is wired to the PC directly (e.g. using a cross-over cable or a local hub), separated from the local area network, the user may simply keep the default parameters to establish the TCP/IP connection.

NOTE

In the **Using Default** mode, the parameters stored in the memory of the module are not cleared automatically. If not changed by the user, they are still available, when switching back to the mode Using Stored.

Link configuration selection


The LAN interface supports 10 or 100 Mbps operation in full- or half-duplex modes. In most cases, full-duplex is supported when the connecting network device - such as a network switch or hub - supports IEEE 802.3u auto-negotiation specifications.

When connecting to network devices that do not support auto-negotiation, the LAN interface will configure itself for 10- or 100-Mbps half-duplex operation.

For example, when connected to a non-negotiating 10-Mbps hub, the LAN interface will be automatically set to operate at 10-Mbps half-duplex.

If the module is not able to connect to the network through auto-negotiation, you can manually set the link operating mode using link configuration switches on the module.

Table 18 Link Configuration Switches

	SW 3	SW 4	SW 5	Link Configuration
	OFF	-	-	speed and duplex mode determined by auto-negotiation
	ON	OFF	OFF	manually set to 10 Mbps, half-duplex
	ON	OFF	ON	manually set to 10 Mbps, full-duplex
	ON	ON	OFF	manually set to 100 Mbps, half-duplex
	ON	ON	ON	manually set to 100 Mbps, full-duplex

Automatic Configuration with BootP

NOTE

All examples shown in this chapter will not work in your environment. You need your own IP-, Subnet-Mask- and Gateway addresses.

NOTE

Assure that the detector configuration switch is set properly. The setting should be either **BootP** or **BootP & Store**, see [Table 16](#) on page 69.

NOTE

Assure that the detector connected to the network is powered off.

NOTE

If the Agilent BootP Service program is not already installed on your PC, then install it from your Agilent ChemStation DVD, located in folder **BootP**.

About Agilent BootP Service

The Agilent BootP Service is used to assign the LAN Interface with an IP address.

The Agilent BootP Service is provided on the ChemStation DVD. The Agilent BootP Service is installed on a server or PC on the LAN to provide central administration of IP addresses for Agilent instruments on a LAN. The BootP service must be running TCP/IP network protocol and cannot run a DHCP server.

How BootP Service Works

When an instrument is powered on, an LAN Interface in the instrument broadcasts a request for an IP address or host name and provides its hardware MAC address as an identifier. The Agilent BootP Service answers this request and passes a previously defined IP address and host name associated with the hardware MAC address to the requesting instrument.

The instrument receives its IP address and host name and maintains the IP address as long as it is powered on. Powering down the instrument causes it to lose its IP address, so the Agilent BootP Service must be running every time the instrument powers up. If the Agilent BootP Service runs in the background, the instrument will receive its IP address on power-up.

The Agilent LAN Interface can be set to store the IP address and will not lose the IP address if power cycled.

Situation: Cannot Establish LAN Communication

If a LAN communication with BootP service cannot be established, check the following on the PC:

- Is the BootP service started? During installation of BootP, the service is not started automatically.
- Does the Firewall block the BootP service? Add the BootP service as an exception.
- Is the LAN Interface using the BootP-mode instead of "Using Stored" or "Using Default" modes?

Installation of BootP Service

Before installing and configuring the Agilent BootP Service, be sure to have the IP addresses of the computer and instruments on hand.

- 1** Log on as Administrator or other user with Administrator privileges.
- 2** Close all Windows programs.
- 3** Insert the Agilent ChemStation software DVD into the drive. If the setup program starts automatically, click **Cancel** to stop it.
- 4** Open Windows Explorer.
- 5** Go to the BootP directory on the Agilent ChemStation DVD and double-click **BootPPackage.msi**.
- 6** If necessary, click the **Agilent BootP Service...** icon in the task bar.
- 7** The **Welcome** screen of the **Agilent BootP Service Setup Wizard** appears. Click **Next**.
- 8** The **End-User License Agreement** screen appears. Read the terms, indicate acceptance, then click **Next**.
- 9** The **Destination Folder** selection screen appears. Install BootP to the default folder or click **Browse** to choose another location. Click **Next**.
The default location for installation is:
C:\Program Files\Agilent\BootPService\
10 Click **Install** to begin installation.

11 Files load; when finished, the **BootP Settings** screen appears.




Figure 27 BootP Settings screen

12 In the **Default Settings** part of the screen, if known, you can enter the subnet mask and gateway.

Defaults can be used:

- The default subnet mask is 255.255.255.0.
- The default gateway is 10.1.1.101.

13 On the **BootP Settings** screen, click **OK**. The **Agilent BootP Service Setup** screen indicates completion.

14 Click **Finish** to exit the **Agilent BootP Service Setup** screen.

15 Remove the DVD from the drive.

This completes installation.

16 Start the BootP service. On the Windows® desktop, select **Start > Control Panel > Services**. Select the **Agilent BootP Service** and click **Start**.

Two Methods to Determine the MAC Address

Enabling logging to discover the MAC address using BootP

If you want to see the MAC address, select the **Do you want to log BootP requests?** check box.

- 1 Open BootP Settings from **Start > All Programs > Agilent BootP Service > EditBootPSettings**.
- 2 In **BootP Settings...** check **Do you want to log BootP requests?** to enable logging.




Figure 28 Enable BootP logging

The log file is located in

C:\Documents and Settings\All Users\Application Data\Agilent\BootP\LogFile

It contains a MAC address entry for each device that requests configuration information from BootP.

- 3 Click **OK** to save the values or **Cancel** to discard them. The editing ends.
- 4 After each modification of the BootP settings (i.e. **EditBootPSettings**) a stop or start of the BootP service is required for the BootP service to accept changes. See [“Stopping the Agilent BootP Service”](#) on page 82 or [“Restarting the Agilent BootP Service”](#) on page 83.
- 5 Uncheck the **Do you want to log BootP requests?** box after configuring instruments; otherwise, the log file will quickly fill up disk space.

Determining the MAC address directly from the LAN Interface card label

- 1 Turn off the instrument.
- 2 Read the MAC address from the label and record it.
The MAC address is printed on a label on the rear of the module. It is the number below the barcode and after the colon (:) and usually begins with the letters AD.
See [Figure 20](#) on page 66 and [Figure 21](#) on page 66.
- 3 Turn on the instrument.

Assigning IP Addresses Using the Agilent BootP Service

The Agilent BootP Service assigns the Hardware MAC address of the instrument to an IP address.

Determining the MAC address of the instrument using BootP Service

- 1 Power cycle the Instrument.
- 2 After the instrument completes self-test, open the log file of the BootP Service using Notepad.
 - The default location for the logfile is C:\Documents and Settings\All Users\Application Data\Agilent\BootP\LogFile.
 - The logfile will not be updated if it is open.The contents will be similar to the following:

02/25/10 15:30:49 PM

Status: BootP Request received at outermost layer

Status: BootP Request received from hardware address: 0010835675AC

Error: Hardware address not found in BootPTAB: 0010835675AC

Status: BootP Request finished processing at outermost layer

- 3 Record the hardware (MAC) address (for example, 0010835675AC).

4 LAN Configuration

Automatic Configuration with BootP


- 4 The Error means the MAC address has not been assigned an IP address and the Tab File does not have this entry. The MAC address is saved to the Tab File when an IP address is assigned.
- 5 Close the log file before turning on another instrument.
- 6 Uncheck the **Do you want to log BootP requests?** box after configuring instruments to avoid having the logfile use up excessive disk space.

Adding each instrument to the network using BootP

- 1 Follow **Start > All Programs > Agilent BootP Service** and select **Edit BootP Settings**. The BootP Settings screen appears.
- 2 Uncheck the **Do you want to log BootP requests?** once all instruments have been added.

The **Do you want to log BootP requests?** box must be unchecked when you have finished configuring instruments; otherwise, the log file will quickly fill up disk space.

- 3 Click **Edit BootP Addresses...** The **Edit BootP Addresses** screen appears.
- 4 Click **Add...** The **Add BootP Entry** screen appears.



The screenshot shows the 'Add BootP Entry' dialog box. It has a title bar with the text 'Add BootP Entry' and a close button (X). The dialog is divided into two main sections. The first section contains three input fields: 'Mac Address', 'Host Name', and 'IP Address' (which has three dots as a placeholder). The second section contains three input fields: 'Comment', 'Subnet Mask' (which has the text '255 . 255 . 255 . 0' entered), and 'Gateway' (which has three dots as a placeholder). At the bottom of the dialog are three buttons: 'OK', 'Cancel', and 'Help'.

Figure 29 Enable BootP logging

5 Make these entries for the instrument:

- MAC address
- Host name, Enter a Hostname of your choice.
The Host Name must begin with "alpha" characters (i.e. LC1260)
- IP address
- Comment (optional)
- Subnet mask
- Gateway address (optional)

The configuration information entered is saved in the Tab File.

6 Click **OK**.

7 Leave **Edit BootP Addresses** by pressing **Close**.

8 Exit **BootP Settings** by pressing **OK**.

9 After each modification of the BootP settings (i.e. EditBootPSettings) a stop or start of the BootP service is required for the BootP service to accept changes. See [“Stopping the Agilent BootP Service”](#) on page 82 or [“Restarting the Agilent BootP Service”](#) on page 83.

10 Power cycle the Instrument.

OR

If you changed the IP address, power cycle the instrument for the changes to take effect.

11 Use the PING utility to verify connectivity by opening a command window and typing:

Ping 10.1.1.101 for example.

The Tab File is located at

C:\Documents and Settings\All Users\Application Data\Agilent\BootP\TabFile

Changing the IP Address of an Instrument Using the Agilent BootP Service

Agilent BootP Service starts automatically when your PC reboots. To change Agilent BootP Service settings, you must stop the service, make the changes, and then restart the service.

Stopping the Agilent BootP Service

- 1 From the Windows control panel, select **Administrative Tools > Services**. The **Services** screen appears.




Figure 30 Windows Services screen

- 2 Right-click **Agilent BootP Service**.
- 3 Select **Stop**.
- 4 Close the **Services and Administrative Tools** screen.

Editing the IP address and other parameters in EditBootPSettings

- 1 Select **Start > All Programs > Agilent BootP Service** and select **Edit BootP Settings**. The **BootP Settings** screen appears.
- 2 When the **BootP Settings** screen is first opened, it shows the default settings from installation.

- 3 Press **Edit BootP Addresses...** to edit the Tab File.




Figure 31 Edit BootP Addresses screen

- 4 In the **Edit BootP Addresses...** screen press **Add...** to create a new entry or select an existing line from the table and press **Modify...** or **Delete** to change the IP address, comment, subnet mask, for example, in the Tab File.
If you change the IP address, it will be necessary to power cycle the instrument for the changes to take effect.
- 5 Leave **Edit BootP Addresses...** by pressing **Close**.
- 6 Exit BootP Settings by pressing OK.

Restarting the Agilent BootP Service

- 1 In the Windows control panel, select **Administrative Tools > Services**. The **Services** screen appears, see [Figure 30](#) on page 82.
- 2 Right-click **Agilent BootP Service** and select **Start**.
- 3 Close the **Services and Administrative Tools** screens.

Manual Configuration

Manual configuration only alters the set of parameters stored in the non-volatile memory of the module. It never affects the currently active parameters. Therefore, manual configuration can be done at any time. A power cycle is mandatory to make the stored parameters become the active parameters, given that the initialization mode selection switches are allowing it.




Figure 32 Manual Configuration (Principle)

With Telnet

Whenever a TCP/IP connection to the module is possible (TCP/IP parameters set by any method), the parameters may be altered by opening a Telnet session.

- 1 Open the system (DOS) prompt window by clicking on Windows **START** button and select **"Run..."**. Type "cmd" and press OK.
- 2 Type the following at the system (DOS) prompt:
 - **c:\>telnet <IP address>** or
 - **c:\>telnet <host name>**




Figure 33 Telnet - Starting a session

where <IP address> may be the assigned address from a Bootp cycle, a configuration session with the Handheld Controller, or the default IP address (see ["Configuration Switch"](#) on page 68).

When the connection was established successfully, the module responds with the following:




Figure 34 A connection to the module is made

- 3 Type **?** and press enter to see the available commands.



Figure 35 Telnet Commands

Table 19 Telnet Commands

Value	Description
?	displays syntax and descriptions of commands
/	displays current LAN settings
ip <x.x.x.x>	sets new ip address
sm <x.x.x.x>	sets new subnet mask
gw <x.x.x.x>	sets new default gateway
exit	exits shell and saves all changes

4 To change a parameter follows the style:

- parameter value, for example:
ip 134.40.28.56

Then press [Enter], where parameter refers to the configuration parameter you are defining, and value refers to the definitions you are assigning to that parameter. Each parameter entry is followed by a carriage return.

- 5 Use the “/” and press Enter to list the current settings.

```

Telnet 134.40.30.205
>/
LAN Status Page
-----
MAC Address   : 0030D317521C
Init Mode     : Using Stored
TCP/IP Properties
- active -
IP Address    : 134.40.30.205
Subnet Mask   : 255.255.248.0
Def. Gateway  : 134.40.24.1
TCP/IP Status : Ready
Controllers   : no connections
>_

```

Figure 36 Telnet - Current settings in "Using Stored" mode

information about the LAN interface
MAC address, initialization mode
Initialization mode is Using Stored
active TCP/IP settings

TCP/IP status - here ready
connected to PC with controller software (e.g. Agilent
ChemStation), here not connected

- 6 Change the IP address (in this example 192.168.254.12) and type “/” to list current settings.

```

Telnet 134.40.30.205
>ip 192.168.254.12
>/
LAN Status Page
-----
MAC Address   : 0030D317521C
Init Mode     : Using Stored
TCP/IP Properties
- active -
IP Address    : 134.40.30.205
Subnet Mask   : 255.255.248.0
Def. Gateway  : 134.40.24.1
- stored -
IP Address    : 192.168.254.12
Subnet Mask   : 255.255.248.0
Def. Gateway  : 134.40.24.1
TCP/IP Status : Ready
Controllers   : no connections
>_

```

Figure 37 Telnet - Change IP settings

change of IP setting to
Initialization mode is Using Stored

active TCP/IP settings


stored TCP/IP settings in non-volatile memory

connected to PC with controller software (e.g. Agilent
ChemStation), here not connected

4 LAN Configuration

Manual Configuration

- 7 When you have finished typing the configuration parameters, type **exit** and press **Enter** to exit with storing parameters.



```
C:\WINDOWS\system32\cmd.exe
Agilent Technologies G4212A PR00100015
>exit

Connection to host lost.
C:\>_
```

Figure 38 Closing the Telnet Session

NOTE

If the Initialization Mode Switch is changed now to “Using Stored” mode, the instrument will take the stored settings when the module is re-booted. In the example above it would be 192.168.254.12.

With the Instant Pilot (G4208A)

To configure the TCP/IP parameters before connecting the module to the network, the Instant Pilot (G4208A) can be used.

- 1 From the Welcome screen press the **More** button.
- 2 Select **Configure**.
- 3 Press the **DAD** button.
- 4 Scroll down to the LAN settings.

Setting	Value
Symbolic Name	<Not Set>
UV-Lamp Tag	Use lamp tag settings
Analog Out 1	0V - 1V output range
UV lamp	Stays off at power on
LAN IP	134. 40. 28. 56
LAN Subnet Mask	255.255.255.0
LAN Def. Gateway	134.40.24.1

Figure 39 Instant Pilot - LAN Configuration (Edit mode)


- 5 Press the **Edit** button (only visible if not in Edit mode), perform the required changes and press the **Done** button.
- 6 Leave the screen by clicking **Exit**.

PC and Agilent ChemStation Setup


PC Setup for Local Configuration

This procedure describes the change of the TCP/IP settings on your PC to match the module's default parameters in a local configuration (see [Table 17](#) on page 72).

- 1 Open the Local Area Connection Properties and select **Internet Protocol (TCP/IP)**. Then click on **Properties**.




- 2 You may enter here the fixed IP address of the module or use the **Alternative Configuration**.



4 LAN Configuration

PC and Agilent ChemStation Setup

- 3 We will use the direct LAN access via Cross-over LAN cable with the module's IP address.



- 4 Click on **OK** to save the configuration.


Agilent ChemStation Setup

This procedure describes the Agilent ChemStation B.04.02 setup for the 1290 Infinity system using the 1290 Infinity DAD (G4212A) as the interfacing module.

NOTE

The LAN must be connected to detector due to high data load on communication to Control Software.

- 1 Open the ChemStation Configuration Editor.




- 2 Select from the menu **Configure - Instruments**.
- 3 Select **Modular 3D LC System**.
- 4 Give the Instrument a name.

4 LAN Configuration


PC and Agilent ChemStation Setup

5 Click on **OK**.



The "Select Instrument" dialog box is shown. It has a title bar "Select Instrument". Below the title bar, there is a label "Instrument Type:" followed by a list box containing the following items: "LC 3D Data Analysis only", "LC Data Analysis only", "Modular 3D LC System" (which is selected), "Modular 3D LC System (Classic)", "Modular LC System", and "Modular LC System (Classic)". Below the list box is a label "Instrument Name:" followed by a text box containing "G4212A". Below the text box is a label "Initially Start Instrument Session?" followed by two radio buttons: "Yes" and "No". Below the radio buttons is a label "Initial Screen Window Size:" followed by three radio buttons: "Normal" (which is selected), "Icon", and "Full screen". At the bottom of the dialog box are three buttons: "OK", "Cancel", and "Help".

6 Select **LC System Access — Access Point** and click on **Add**.




The "G4212A - Device Configuration" dialog box is shown. It has a title bar "G4212A - Device Configuration". Below the title bar, there is a label "Modular 3D LC System". Below the label is a list box labeled "Modules" containing the following items: "LC System Access", "Access Point" (which is selected), "Dual Channel Iface", "35900E", "CTCAutoSampler", and "CTC PAL AutoSampler". To the right of the list box are two radio buttons: "GPIB" (which is selected) and "LAN". Below the radio buttons is a label "GPIB Address:" followed by a text box and a "GPIB Address" button. Below the list box is a label "Selected Modules" followed by a list box containing the following items: "LC System Access", "Access Point", and an empty space. To the right of the list box is a "Delete" button. At the bottom of the dialog box are three buttons: "OK", "Cancel", and "Help".

7 Click on **OK**.

The Configuration Editor shows now the new instrument.

8 If required, change under **Configure – Path** the folder locations.

9 Save the current configuration via **File – Save**.



10 Exit the Configuration Editor.


11 Start the Agilent ChemStation.

During first startup or when the system configuration has changed, a notification shows up.


4 LAN Configuration

PC and Agilent ChemStation Setup

- 12** The left column shows the modules that could be configured. You may select the module manually from the list. We use the Auto Configuration mode. Click on **Yes**.




- 13** Enter the IP address or the Hostname of the module with the LAN-access.




- 14** Click on **OK**.

The selected module is shown now in the right window (with serial number). In addition all other modules connected via CAN to the detector are shown as well.

15 Click on **OK** to continue the ChemStation loading.



16 You may see the details of the module by **selecting the module** and clicking on **Configure**.



Under **Connection Settings** you may change the IP/Hostname of the module (may require a re-start of the ChemStation).

After successful load of the ChemStation, you should see the module(s) as active item in the graphical user interface (GUI).

4 LAN Configuration

PC and Agilent ChemStation Setup





Figure 40 Screen After Successful Load of ChemStation



5 Using the Module

Preparing the Detector	100
Setting up the Detector with Agilent ChemStation	101
The Detector GUI	103
Control Settings	106
Method Parameter Settings	107
General Method Settings	108
Advanced Method Parameter Settings	110
Spectrum Settings	111
Other Advanced Method Parameter Settings	113
Time Table Settings	114
Instrument Curves	116
Instrument Configuration	117
Main Screens of the Detector with Agilent Instant Pilot (G4208A)	119

This chapter provides information on how to set up the module for an analysis and explains the basic settings.



Preparing the Detector

For best performance of the detector

- Let the lamp warm-up and stabilize for at least one hour (initial turn on of the module requires a longer time depending on the environment and the application needs); refer to “[Specification Conditions](#)” on page 43.
- For high sensitivity measurements, a stable environment is required; refer to “[Environment](#)” on page 39. Prevent drafts from air condition systems.
- Setting an appropriate reference wavelength could improve the baseline behavior. Alternatively use the 1.6 µL heat exchanger from the G1316C TCC.
- Do not work with removed front panels. When the front panel of the G1316C TCC (typically located below the detector) is removed while the TCC is set to high temperatures, the up-streaming air could influence the stability of the detector baseline.

Setting up the Detector with Agilent ChemStation

The setup of the detector is shown with the Agilent ChemStation B.04.02 based on the 1290 Infinity DAD (G4212A). Depending on the controller (e.g. Agilent Instant Pilot, EZChrom Elite, MassHunter) the screens look different. For the Instant Pilot refer to [“Main Screens of the Detector with Agilent Instant Pilot \(G4208A\)”](#) on page 119.

NOTE

This section describes the detector settings only. For information of the Agilent ChemStation or other Agilent 1260 Infinity/1290 Infinity Series modules refer to the corresponding documentation or system manual.

After successful load of the ChemStation, you should see the module as an active item in the graphical user interface (GUI).

5 Using the Module

Setting up the Detector with Agilent ChemStation




Figure 41 ChemStation Method & Run

The Detector GUI



Within the detector GUI, there are active areas. If you move the mouse cursor across the icons the cursor will change and you may click on the button (1) to

- “Make Device Ready/Turn device off (standby)”
- turn on/off the lamp



Signal Information, activated by button (1), shows actual values of all selected signals


- Signal name (A, B, C, ...)
- Sample wavelength/bandwidth
- Reference wavelength/bandwidth
- Absorbance

If more signals are activated, the size of the detector GUI will change accordingly.

	WL	BW	RefWL	RefBW [mAU]	
A	254.0	4	360.0	100	1.7
B	210.0	4	360.0	100	14.1
C	214.0	4	360.0	100	13.8
D	230.0	4	360.0	100	1.1
E	260.0	4	360.0	100	1.2

5 Using the Module

Setting up the Detector with Agilent ChemStation




- A right-click into the *Active Area* will open a menu to
- Show the Control Interface (special module settings)
 - Show the Method interface (similar as via menu Instrument – Setup Instrument Method)
 - Set Error Method
 - Identify Module (Status LED will blink)
 - Perform a Balance
 - Switch the UV-lamp on/off (same as click on button “Make Device Ready/Turn device off (standby)”)



Module Status shows Run / Ready / Error state and “Not Ready text” or “Error text”.


- Error (Red)
- Not ready (yellow)
- Ready (green)
- Pre run, Post run (purple)
- Run (blue)
- Idle (green)
- Offline (dark gray)
- Standby (light gray)



EMF Status shows

- Offline (gray)
- Ok. No Maintenance required
- EMF warning. Maintenance or check might be required (yellow)
- EMF warning. Maintenance required (red)

Important: The EMF settings can be accessed via the Agilent Lab Advisor or the Instant Pilot only. The limit(s) can be changed. Based on the limit, the User Interface displays the above status.



RFID tag information is displayed when moving with the mouse cursor on to the tag attached to the flow cell or lamp. The information provides flow cell and lamp related information like

- Part number
 - Production date
 - Serial number
- and other details.

Lamp tag information

Burn time	93.3 h
Minimum lifetime	2000.0 h
Number of ignitions	10
Product Number	5190-0917
Serial Number	824337
Production Date	4/9/2009 8:23:53 AM
Tested Date	7/16/2009 1:50:04 PM
Intensity at test	37275 counts

Cell tag information

Cell Name	Max-Light Cell
Product Number	G4212-60008
Serial Number	10PP042325
Production Date	2/5/2009 12:49:06 PM
Optical path length	10.0 mm
Cell Volume (σ)	1.0 μ L
Maximum pressure	60 bar
Tested Date	7/10/2009 1:44:52 PM
Cell Revision	0

5 Using the Module

Setting up the Detector with Agilent ChemStation

Control Settings



Lamps: can be turned ON/OFF.

At Power On: automatic lamp-on at power on.

Analog Output Range: can be set to either 100 mV or 1 V full scale (1 V = default).

UV lamp Tag: Automatic detects a lamp with RFID tag. If no RFID tag lamp is used, “*UV lamp not ready*” is displayed and it cannot be ignited. A compatible mode has to be selected based on the used lamp; see Non-RFID-tag lamp information below.

Automatic Turn On: The module can be turned on at a specified date/time. If “*Turn UV lamp on*” at power on is set, the lamp is turned on as well.

Non-RFID-tag lamp

In case a non-RFID-tag lamp is used, the user interface will show this when selecting a compatible mode.

You may operate the detector outside of the guaranteed specification.

Method Parameter Settings

These settings are available via **Menu – Instrument – Setup Instrument Method** or via right click on the active area of the detector GUI.

NOTE


The Instrument Curves tab is not shown when opening the method parameter settings via right mouse click on the detector GUI.

Switches to additional instrument signals for troubleshooting

Toggles to Advanced Settings (actual window)

Toggles to Timetable Settings

Opens the Time table graph



General Method Settings

Signals

	Use Signal	Wave length	Band width	Reference Wavelength	Reference Bandwidth	
Signal A	<input checked="" type="checkbox"/>	254.0	4.0	<input checked="" type="checkbox"/> 360.0	100.0	nm
Signal B	<input checked="" type="checkbox"/>	210.0	4.0	<input checked="" type="checkbox"/> 360.0	100.0	nm
Signal C	<input checked="" type="checkbox"/>	214.0	4.0	<input checked="" type="checkbox"/> 360.0	100.0	nm
Signal D	<input checked="" type="checkbox"/>	230.0	4.0	<input checked="" type="checkbox"/> 360.0	100.0	nm
Signal E	<input checked="" type="checkbox"/>	260.0	4.0	<input checked="" type="checkbox"/> 360.0	100.0	nm
Signal F	<input type="checkbox"/>	273.0	4.0	<input checked="" type="checkbox"/> 360.0	100.0	nm
Signal G	<input type="checkbox"/>	280.0	4.0	<input checked="" type="checkbox"/> 360.0	100.0	nm
Signal H	<input type="checkbox"/>	250.0	100.0	<input checked="" type="checkbox"/> 360.0	100.0	nm

Up to 8 individual signals can be set. For each of the signals, the wavelength and bandwidth can be set for sample and reference.

Limits:

Wavelength: 190.0 to 640.0 nm in steps of 0.1 nm

Bandwidth: 1.0 to 400.0 nm in steps of 0.1 nm

Setting an appropriate reference wavelength could improve the baseline behavior. Alternatively use the 1.6 µl heat exchanger from the G1316C TCC or an optional DAD heat exchanger (if available).

Peakwidth

Peakwidth

>0.10 min (2.0 s response time) (2.5 Hz)

Stoptime

As

<0.0016 min (0.016 s response time) (160 Hz)

>0.0016 min (0.03 s response time) (160 Hz)

>0.003 min (0.062 s response time) (80 Hz)

>0.006 min (0.12 s response time) (40 Hz)

>0.012 min (0.25 s response time) (20 Hz)

>0.025 min (0.5 s response time) (10 Hz)

>0.05 min (1.0 s response time) (5 Hz)

>0.10 min (2.0 s response time) (2.5 Hz)

>0.20 min (4.0 s response time) (1.25 Hz)

>0.40 min (8.0 s response time) (0.62 Hz)

>0.85 min (16.0 s response time) (0.31 Hz)

min

Peakwidth enables you to select the peak width (response time) for your analysis. The peak width is defined as the width of a peak, in minutes, at half the peak height. Set the peak width to the narrowest expected peak in your chromatogram. The peak width sets the optimum response time for your detector. The peak detector ignores any peaks that are considerably narrower, or wider, than the peak width setting. The response time is the time between 10 % and 90 % of the output signal in response to an input step function. When the All spectrum storage option is selected, then spectra are acquired continuously depending on the setting of the peak width. The time specified by the peak width is used as a factor in the acquisition of spectra. The acquisition time for one spectrum is slightly less than the peak width divided by 8, which is the acquisition time.

Limits: When you set the peak width (in minutes), the corresponding response time is set automatically and the appropriate data rate for signal and spectra acquisition is selected.

- Do not use peak width shorter than necessary.
- Do not use 0.025 seconds response time (no filtering/high noise and no need (actually ultra-fast LC doesn't deliver peaks < 0.0025 min/ < 0.15 sec).

NOTE

The 1260 Infinity DAD (G4212B) has a data rate of up to 80 Hz.

Peakwidth (time programmed)

The screenshot shows a dropdown menu titled "Peakwidth". The menu is open, displaying four options: "same as Peakwidth" (highlighted in orange), "2-times greater than Peakwidth", "4-times greater than Peakwidth", and "8-times greater than Peakwidth".

These selections can be made during time programmed operation.

When used in a timetable, Peakwidth changes the filters used for peak-controlled spectra acquisition, but not the data rate of a chromatographic signal.

NOTE

This setting makes sense only with peak-controlled spectra; it allows you to change the peakwidth setting to account for broadening peaks at the end of the run.

Stoptime / Posttime

The screenshot shows two side-by-side panels. The left panel, titled "Stoptime", has two radio buttons: "As Pump/Injector" (selected) and a numeric input field set to "1.00" with a "min" unit label. The right panel, titled "Posttime", has two radio buttons: "Off" (selected) and a numeric input field set to "1.00" with a "min" unit label.

The stoptime is the time where either the complete system stops (As Pump/Injector) or the module (if different from system stop time). The data collection is stopped at this time.

A posttime period can be used to allow module's items to equilibrate (e.g. after gradient change or temperature change).

Advanced Method Parameter Settings

These settings are available via a click on the **Advanced** link of the Method Parameter Settings (in case the Time Table Settings are open).

This screen shows the default settings.

The screenshot shows the 'Advanced' settings window for a detector. It is organized into several sections:

- Spectrum**: Contains a 'Store' dropdown set to 'None', a 'Range from' field set to 190.0 nm, a 'to' field set to 400.0 nm, a 'Step' field set to 2.0 nm, and a 'Threshold' field set to 10.0 mAU.
- Analog Output**: Contains 'Output 1' settings with a 'Zero Offset' field set to 5 % and an 'Attenuation' dropdown set to 1000 mAU.
- Margin for negative Absorbance**: A field set to 100 mAU.
- Slit**: A dropdown menu set to 4 nm.
- Autobalance**: Two checkboxes, 'Prerun' (checked) and 'Postrun' (unchecked).
- Lamps on required for acquisition**: A checkbox for 'UV Lamp' which is checked.

Figure 42 Method Parameter Settings

NOTE

The 1260 Infinity DAD (G4212B) has a fixed slit width of 4 nm.

Spectrum Settings

Store

The screenshot shows the 'Spectrum' settings window. Under the 'Store' tab, a dropdown menu is open, listing various acquisition points: None, Apex, Apex+Baselines, Apex+Slopes, Apex+Slope+Baselines, All in Peak, Every 2nd Spectrum, and All. The 'Range from' field is set to 'Apex' and the 'Threshold' field is set to 'All in Peak'. Below this, the 'Analog Output' section shows 'Output 1' set to 'All'.

Defines at which points on “*signal A*” spectra will be taken and saved. Signal A is used to control the “*peak controlled spectra acquisition*”; the other signals have no influence on spectra acquisition.

Limits:

190.0 to 640.0 nm in steps of 0.1 nm for both low and high values. The high value must be greater than the low value by at least 0.1 nm.

None

No spectra are taken.

Apex

Spectra are taken at the apex of the peak.

Apex + Baselines

Spectra are taken at the apex and baselines of the peak.

Apex + Slopes

Spectra are taken at the apex, upslope, and down slope of the peak.

Apex + Slopes + Baselines

Spectra are taken at the apex, baselines, upslope, and down slope of the peak.

All in Peak

All spectra within the peak are taken.

Every 2nd spectrum

Spectra are taken continuously as for All, but only every second spectrum is stored; other spectra are discarded. This reduces the amount of data storage necessary.

Range

Range defines the wavelength range for spectral storage. Limits: 190 to 650 nm in steps of 1 nm for both low and high values. The high value must be greater than the low value by at least 2 nm.

5 Using the Module

Setting up the Detector with Agilent ChemStation

Step	Step defines the wavelength resolution for spectral storage. Limits: 0.10 to 100.00 nm nm in steps of 0.1 nm.
Threshold	The threshold is the height in mAU of the smallest expected peak. The peak detector ignores any peaks which are lower than the threshold value and does not save spectra. Limits: 0.001 to 1000.00 mAU in steps of 0.001 mAU.

Other Advanced Method Parameter Settings

The screenshot shows the 'Analog Output' configuration window. It includes fields for 'Output 1:', 'Zero Offset' (5%), 'Attenuation' (1000 mAU), 'Margin for negative Absorbance' (100 mAU), 'Slit' (4 nm), 'Autobalance' (Prenun checked, Postrun unchecked), and 'Lamps on required for acquisition' (UV Lamp checked).

This screen (part of the Advanced Method Settings) shows the default settings.

Analog Output

The Range can be set to either 100 mV or 1 V full scale, see [“Control Settings”](#) on page 106.

Zero Offset

1 to 99 % in steps of 1 % (5 % equal to 50 mV).

Attenuation

0.98 to 2000 mAU at discrete values for either 100 mV or 1 V full scale.

Margin for Negative Absorbance

Use this field to modify the detector's signal handling to increase the margin for negative absorbance. Use this option if, for example, your solvent gradient produces a decreasing baseline absorbance, and for GPC analyses. Limits: 100 to 4000 mAU.

The higher the value the greater the baseline noise. Set this value only if you expect negative absorbance greater than -100 mAU.

Slit (G4212A)

You can select the optical bandwidth (1, 2, 4 or 8 nm) of the detector; the narrower the slit, the smaller the optical bandwidth of the instrument, but the lower its sensitivity. The smaller the optical bandwidth the higher the spectral resolution.

Autobalance


Defines, whether a balance is performed prior to a run and/or after a run has finished.

Lamp on required for analysis

If unchecked, the lamp will be turned off after the analysis has finished.

Time Table Settings

Timetable window




You may set up time events to change functions with their parameters over the run time. Add lines as required.

Time Limits:

0.00 to 99999.00 minutes in steps of 0.01 min.

Via the buttons in the bottom area, time table lines can be added, removed, cut copied, pasted or completely cleared.

Functions




You may set up time events to change functions with their parameters over the run time. Add lines as required.

Limits:

0.00 to 99999.00 minutes in steps of 0.01 min.


Parameter



Based on the chosen function, a certain parameter can be selected.

Time Table Graph


In this view the activated signals are shown how they change according the time table.



5 Using the Module

Setting up the Detector with Agilent ChemStation

Instrument Curves



The detector has several signals (internal temperatures, voltages of lamps) that can be used for diagnosing problems. These can be baseline problems deriving from deuterium lamps wander / drift problems due to temperature changes.

These signals can be used in addition to the normal baseline signal to determine whether correlation to temperature or voltage/current of the lamp.

These signals are available via the Agilent ChemStation Online Plot/Data Signal and/or Agilent Lab Advisor Software.

Instrument Configuration

These settings are available via menu **Instrument – Instrument Configuration**.





Figure 43 Menu Instrument Configuration

Via the **Instrument Configuration** screen additional modules can be added to a system.


5 Using the Module

Setting up the Detector with Agilent ChemStation



The dialog box titled "Agilent 1100/1200/1290 Series LC Auto Configurati..." contains two radio buttons. The first radio button, labeled "IP address", is selected and is followed by a text field containing "192.168.254. 11". The second radio button is labeled "Hostname" and is followed by an empty text field. At the bottom of the dialog are three buttons: "OK", "Cancel", and "Help".

Use the **Auto Configuration** to define the LAN communication between the Agilent ChemStation and the host module (usually the Agilent detector). Changing parameters become active after reboot of the ChemStation.



The dialog box titled "1100/1200/1290 DAD Configuration: Instrument 1" has a "Communication" tab. Under this tab, there are four fields: "Device name" with the value "1290 Infinity DAD", "Type ID" with a dropdown menu showing "G4212A", "Serial number" with the value "PR00100015", and "Firmware revision" with the value "B.06.23 [0003]". Below these fields is a button labeled "Connection settings...". At the bottom of the dialog is an "Options" section with an empty list box. At the very bottom are three buttons: "OK", "Cancel", and "Help".

Device name: based on the module.

Type ID: based on the module (product number). Some modules may allow changing the type based on hardware/firmware. This results in a change of features and functions.


Serial number: based on the module.

Firmware revision: based on the module.


Options: lists installed options.

Main Screens of the Detector with Agilent Instant Pilot (G4208A)

Below the main screens for the use of the detector are shown.



The Welcome screen shows all modules of the system.




The **Control** screen allows

- Lamp On/Off
- Get Ready
- Reset of Errors
- Balance
- Take Spectrum

5 Using the Module

Main Screens of the Detector with Agilent Instant Pilot (G4208A)




The **System Info** screen displays a table of detector properties and their values. On the right side, there are buttons for 'Reload', 'Print', and 'Exit'. At the bottom, there is a status bar with the text 'Information on each module.' and a clock showing '16:43'.

Property	Value
Controller : PP5S055002	
Main Revision	T.00.05 [0003]
Board ID	TYPE=G4208-66500, SER=001077, REV=B.02.00, M
	DAD : PR00100015
Main Revision	B.06.23 [0003]
Resident Revision	B.06.20 [0001]
On-time	0d 03:34h
LAN TCP/IP Mode	USING STORED
LAN TCP/IP Address	192.168.254.11
LAN MAC Address	0030D317521C
Board ID	TYPE=G4212-65800, REV=BE, MFG=ZZ, SER=M/
Lamp	5190-0917 : \$24337
Cell	G4212-60011 : 200PP00015

The **System Info** screen list details of the detector

- Firmware revision
- On-time
- LAN settings
- Main Board information
- Lamp RFID tag Information
- Flow cell RFID tag Information



The **Configure - DAD** screen shows a table of settings. On the right, there are 'Cancel' and 'Done' buttons. At the bottom, there is a text input field with the placeholder 'Enter up to 30 characters' and a clock showing '16:43'.

Setting	Value
Symbolic Name	
UV-Lamp Tag	Use lamp tag settings
Analog Out 1	0V - 1V output range
UV lamp	Stays off at power on
LAN IP	192.168.254.11
LAN Subnet Mask	255.255.255.0
LAN Def. Gateway	255.255.255.255

The **Configuration** screen allows to configure

- Symbolic name of module
- Temperature control
- Lamp and cell RFID tag use
- Analog Output range
- UV lamp at power on
- LAN settings

Main Screens of the Detector with Agilent Instant Pilot (G4208A)

Method - DEFAULT all

Setting	Value
DAD : PP00099010	
Stoptime	OFF
Posttime	OFF
Signal A	254.0, 4.0; Ref. 360.0, 100.0 nm
Signal B	210.0, 4.0; Ref. 360.0, 100.0 nm
Signal C	214.0, 4.0; Ref. 360.0, 100.0 nm
Signal D	230.0, 4.0; Ref. 360.0, 100.0 nm
Signal E	260.0, 4.0; Ref. 360.0, 100.0 nm
Signal F	273.0, 4.0; Ref. 360.0, 100.0 nm
Signal G	280.0, 4.0; Ref. 360.0, 100.0 nm
Signal H	250.0, 100.0; Ref. 360.0, 100.0 nm
UV Lamp Required	Yes
Spectrum	190.0 nm - 400.0 nm step 2.0 nm
Slit Width	4 nm

Cancel


Done

14:51

The **Method** screen lists all method parameters of the detector. These can be edited.

5 Using the Module


Main Screens of the Detector with Agilent Instant Pilot (G4208A)



The **Maintenance** screen allows


- EMF setup
- Maintenance (calibrate, cell/lamp info)
- logging of maintenance activities
- module identification (blinking LED)

Firmware Updates can be done via the System Maintenance screen.



The **Diagnose** screen provides access to module specific tests

- Lamp Intensity
- Calibration
- Cell



6 Optimizing the Detector

Introduction	124
Optimization Overview	125
Choosing a Flow Cell	126
Optimizing for Sensitivity, Selectivity, Linearity and Dispersion	127
Flow Cell Path Length	127
Peak width (response time)	129
Sample and Reference Wavelength and Bandwidth	131
Slit Width (G4212A)	134
Optimizing Spectral Acquisition	136
Margin for Negative Absorbance	137
Optimizing Selectivity	138
Quantifying Coeluting Peaks by Peak Suppression	138
Ratio Qualifiers for Selective Detection of Compound Classes	140
Optimizing the Detector Regarding to the System	142
Delay Volume and Extra-Column Volume	142
How to Configure the Optimum Delay Volume	142
How to Achieve Higher Sensitivity	143
Warm up of the Detector	149

This chapter provides information on how to optimize the detector.



Introduction

The detector has a variety of parameters that can be used to optimize performance. Depending on whether signal or spectral data need to be optimized, different settings are recommended. The following sections describe optimization for:

- signal sensitivity, selectivity and linearity,
- spectral sensitivity and resolution (DAD only), and
- disk space required for storing data.

NOTE

The information in this chapter should be seen as a basic introduction to diode array detector techniques. Some of these techniques may not be available in the instrument software controlling the detector.

How to Get the Best Detector Performance

The information below will guide you on how to get the best detector performance. Follow these rules as a start for new applications. It gives rules-of-thumb for optimizing detector parameters.

Optimization Overview

Table 20 Optimization Overview

Parameter	Impact
1 Selection of flow cell <ul style="list-style-type: none"> Choose flow cell according to used column ("Choosing a Flow Cell" on page 126). 	<ul style="list-style-type: none"> peak resolution versus sensitivity
2 Connection of flow cell	<ul style="list-style-type: none"> chromatographic resolution
3 Setting the peak width (response time) <ul style="list-style-type: none"> Use peak width according "Choosing a Flow Cell" on page 126 as starting point. Set the peak-width close to the width of a narrow peak of interest in your chromatogram. 	<ul style="list-style-type: none"> peak resolution versus sensitivity versus disk space
4 Setting wavelength and bandwidth <ul style="list-style-type: none"> Sample wavelength: <ul style="list-style-type: none"> Never miss a peak by the use of a browser wavelength like 250 nm with 100 nm bandwidth. Select specific wavelength with reduced optical bandwidth if you need selectivity, e.g. 254.0 nm / 4 nm and 360.0 nm / 100 nm as reference wavelength. Set the sample wavelength to a peak or valley to get best linearity in general; select a valley to get best linearity for high concentrations. Reference wavelength: <ul style="list-style-type: none"> Select the reference wavelength with broad bandwidth (30...100 nm) wavelength range where your analytes have little or no absorbance (e.g. sample at 254 nm, reference at 320 nm). Select the reference wavelength as near as possible to the UV range. 	<ul style="list-style-type: none"> sensitivity versus selectivity sensitivity versus linearity baseline drift due to RI effects.
5 Setting the slit width (G4212A only) <ul style="list-style-type: none"> Use 4 nm slit for normal applications. Use narrow slit (e.g 1 nm) if your analytes have narrow absorbance bands and for high concentrations. Use a wide slit (e.g. 8 nm) to detect very low concentrations. Optimizing spectral acquisition (DAD only) Select spectra acquisition mode according to your needs (see "Spectrum Settings" on page 111). Set the spectral wavelength range (for colorless samples 190...400 nm is sufficient). Set step to 4 nm for normal use; set small step (and slit width) if high resolution of spectra with fine structure is wanted. 	<ul style="list-style-type: none"> spectral resolution, sensitivity and linearity.

Choosing a Flow Cell

The Max-Light Cartridge Cell with path length 10 mm and volume (σ) 1.0 μ l covers a wide range of applications:

- all column diameter down to at least 2.1 mm ID or even less
- applications with peak dispersion (Peakwidth x flow) down to $\sim 2 \mu$ l
[example: $\text{pw} = 0.04 \text{ min}$ at $\text{flow} = 0.1 \text{ ml/min}$ gives peak dispersion of $0.04 \text{ min} \times 0.1 \text{ ml/min} = 0.004 \text{ ml} = 4 \mu\text{l}$]

If higher sensitivity is necessary, the Max-Light Cartridge Cell with path length 60 mm and volume (σ) 4 μ L can be used. This cell enhances the detector by lowering the limit of detection (LOD) by a factor of about 3 (depending on the application).

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion

Flow Cell Path Length

Lambert-Beer's law shows a linear relationship between the flow cell path length and absorbance.

$$\text{Absorbance} = -\log T = \log \frac{I_0}{I} = \epsilon \times C \times d$$

where

T is the transmission, defined as the quotient of the intensity of the transmitted light I divided by the intensity of the incident light, I_0 ,

ϵ is the extinction coefficient, which is a characteristic of a given substance under a precisely-defined set of conditions of wavelength, solvent, temperature and other parameters,

C [mol/L] is the concentration of the absorbing species, and

d [cm] is the path length of the cell used for the measurement.

The detector can now output the signal in two forms:

- 1 In Absorbance divide by the path length AU/cm, that is then similar to [$\epsilon \times C$]. Advantage: samples with same concentration have same peak height also at cells with different path lengths.

The upper limit of concentration: the linearity limit of the detector is then seen at about 2 AU/path length, so for the 6 cm Max-Light Cartridge Cell the linearity limit is 333 mAU/cm].

- 2 In AU that is equal to $\epsilon \times C \times d$ like normal done in the past: now for recalculation to your concentration C the path length must be considered.

Therefore, flow cells with longer path lengths yield higher signals. Although noise usually increases little with increasing path length, there is a gain in signal-to-noise ratio.

When increasing the path length, the cell volume could increase. Depending on the peak volume, this could cause more peak dispersion.

6 Optimizing the Detector

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion

As a rule-of-thumb the flow cell volume should be about 1/3 of the peak volume at half height. To determine the volume of your peaks, take the peak width as reported in the integration results multiply it by the flow rate and divide it by 3).

NOTE

This may result in problems when the used peak width is set to large and all peaks are filtered accordingly.

Traditionally LC analysis with UV detectors is based on comparing measurements with internal or external standards. To check photometric accuracy of the Agilent detector it is necessary to have more precise information on path lengths of the detector flow cells.

Part Number	Path Length	Cell Volume (σ)
G4213-60008	1.0 cm	1.0 μ L
G4213-60007	6.0 cm	4.0 μ L

Peak width (response time)

Response time describes how fast the detector signal follows a sudden change of absorbance in the flow cell. The detector uses digital filters to adapt response time to the width of the peaks in your chromatogram. These filters do not affect peak area nor peak symmetry. When set correctly, such filters reduce baseline noise significantly (Figure 44 on page 129), but reduce peak height only slightly. In addition, these filters reduce the data rate to allow optimum integration and display of your peaks and to minimize disk space required to store chromatograms and spectra.




Figure 44 Influence of Response Time on Signal and Noise

Table 21 on page 130 lists the filter choices of the detector. To get optimum results, set peak width as close as possible to a narrow peak of interest in your chromatogram. Response time will be approximately 1/3 of the peak width, resulting in less than 5 % peak-height reduction and less than 5 % additional peak dispersion. Decreasing the peak width setting in the detector will result in less than 5 % gain in peak height but baseline noise will increase by a factor of 1.4 for a factor of 2 response-time reduction. Increasing peak width (response time) by factor of two from the recommended setting (over-filtering) will reduce peak height by about 20 % and reduce baseline

6 Optimizing the Detector

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion

noise by a factor of 1.4. This gives you the best possible signal-to-noise ratio, but may affect peak resolution.

Table 21 Peak Width — Response Time — Data Rate

Peak width at half height [min] ¹	Response [s]	Signal data rate [Hz]	Scan data rate [HZ] ≤126 pts/scan	Scan data rate [HZ] ≤251 pts/scan	Scan data rate [HZ] ≤501 pts/scan	Scan data rate [HZ] >501 pts/scan
< 0.0016	0.016	160 ²	160 ²	80	40	20
> 0.0016	0.03	160 ²	160 ²	80	40	20
> 0.003	0.062	80	80	80	80	40
> 0.006	0.12	40	40	40	40	40
> 0.012	0.25	20	20	20	20	20
> 0.025	0.5	10	10	10	10	10
> 0.05	1.0	5	5	5	5	5
> 0.10	2.0	2.5	2.5	2.5	2.5	2.5
> 0.20	4.0	1.25	1.25	1.25	1.25	1.25
> 0.40	8.0	0.625	0.62	0.625	0.625	0.625
> 0.85	16.0	0.3125	0.31	0.3125	0.3125	0.3125

¹ Values in the User Interface may be rounded.

² G4212A only

NOTE

The maximum spectra scan rate depends on the data points per scan, see [Table 21](#) on page 130. Running at 160 Hz, the spectra scan data rate is reduced automatically if the spectra scan data rate is more than 251 points/scan.

Sample and Reference Wavelength and Bandwidth

The detector measures absorbance simultaneously at wavelengths from 190 to 640 nm. A UV-lamp provides good sensitivity over the whole wavelength range.

If you know little about the analytes in your sample, store all spectra over the full wavelength range. This provides full information but fills up your disk space rather quickly. Spectra can be used to check a peak's purity and identity. Spectral information is also useful to optimize wavelength settings for your chromatographic signal.

The detector can compute and store at run time up to 8 signals with these properties:

- sample wavelength, the center of a wavelength band with the width of sample bandwidth (BW), and optionally
- reference wavelength, the center of a wavelength band with the width of reference bandwidth.

The signals comprises a series of data points over time, with the average absorbance in the sample wavelength band minus the average absorbance of the reference wavelength band.

Signal A in the detector default method is set to sample 254.0/4, reference 360.0/100, that is, the average absorbance from 252 – 256 nm minus the average absorbance from 310 – 410 nm. As all analytes show higher absorbance at 252 – 256 nm than at 310 – 410 nm, this signal will show you virtually every compound which can be detected by UV absorbance.

Many compounds show absorbance bands in the spectrum. [Figure 45](#) on page 132 shows the spectrum of anisic acid as an example. To optimize for lowest possible detectable concentrations of anisic acid, set the sample wavelength to the peak of the absorbance band (that is, 252 nm) and the sample bandwidth to the width of the absorbance band (that is, 30 nm). A reference of 360,100 is adequate. Anisic acid does not absorb in this range.

If you work with high concentrations, you may get better linearity above 1.5 AU by setting the sample wavelength to a valley in the spectrum, like 225 nm for anisic acid.

6 Optimizing the Detector

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion




Figure 45 Optimization of Wavelength Setting

A wide bandwidth has the advantage of reducing noise by averaging over a wavelength range – compared to a 4 nm bandwidth, the baseline noise is reduced by a factor of approximately 2.5, whereas the signal is about 75 % of a 4 nm wide band. The signal-to-noise ratio for a 30 nm bandwidth is twice that for a 4 nm bandwidth in our example.




Figure 46 Influence of Bandwidth on Signal and Noise

Because the detector averages absorbance values that are calculated for each wavelength, using a wide bandwidth does not negatively impact linearity.

The use of a reference wavelength is highly recommended to further reduce baseline drift and wander induced by room temperature fluctuations or refractive index changes during a gradient.

An example of the reduction of baseline drifts is shown in [Figure 47](#) on page 133 for PTH-amino acids. Without a reference wavelength, the chromatogram drifts downwards due to refractive index changes induced by the gradient. This is almost completely eliminated by using a reference wavelength. With this technique, PTH-amino acids can be quantified in the low picomole range even in a gradient analysis.




Figure 47 Gradient Analysis of PTH-Amino Acids (1 pmol each), with and without Reference

Slit Width (G4212A)

The 1290 Infinity DAD (G4212A) has a variable slit at the entrance of the spectrograph. This is an effective tool to adapt the detector to changing demand of different analytical problems.

A narrow slit provides spectral resolution for analytes with very fine structures in the absorbance spectrum. An example of such a spectrum is benzene. The five main absorbance bands (fingers) are only 2.5 nm wide and just 6 nm apart from each other.




Figure 48 Benzene at 1 and 4 nm slit width (principle)

A wide slit uses more of the light shining through the flow cell. This gives lower baseline noise as shown in [Figure 49](#) on page 135.

Slit width = 8 nm

Slit width = 1 nm

Data Rate = 160 Hz




Figure 49 Influence of the Slit Width on Baseline Noise

However, with a wider slit, the spectrograph's optical resolution (its ability to distinguish between different wavelengths) diminishes. Any photodiode receives light within a range of wavelength determined by the slit width. This explains why the fine spectral structure of benzene disappears when using a 8 nm wide slit.

Furthermore, the absorbance is no longer strictly linear with concentration for wavelengths at a steep slope of a compound's spectrum.

Substances with fine structures and steep slopes like benzene are very rare.

In most cases the width of absorbance bands in the spectrum is more like 30 nm as with anisic acid ([Figure 45](#) on page 132).

In most situations, a slit width of 4 nm will give the best results.

Use a narrow slit (1 or 2 nm) if you want to identify compounds with fine spectral structures or if you need to quantify at high concentrations (> 1000 mAU) with a wavelength at the slope of the spectrum. Signals with a wide bandwidth can be used to reduce baseline noise. Because (digital) bandwidth is computed as average of absorbance, there is no impact on linearity.

Use a wide (8 nm) slit when your sample contains very small concentrations. Always use signals with bandwidth at least as wide as the slit width.

Optimizing Spectral Acquisition

Storage of all spectra consumes a lot of disk space. It is very useful to have all spectra available during optimization of a method or when analyzing unique samples. However when running many samples of the same type, the large size of data files with all spectra may become a burden. The detector provides functions to reduce the amount of data, yet retaining the relevant spectral information.

For spectra options see [“Spectrum Settings”](#) on page 111.

Range

Only the wavelength range where the compounds in your sample absorb contains information that is useful for purity checks and library searches. Reducing the spectrum storage range saves disk space.

Step

Most substances have broad absorbance bands. Display of spectra, peak purity and library search works best if a spectrum contains 5 to 10 data points per width of the absorbance bands. For anisic acid (the example used before) a step of 4 nm would be sufficient. However a step of 2 nm gives a more optimal display of the spectrum.

Threshold

Sets the peak detector. Only spectra from peaks higher than threshold will be stored when a peak-controlled storage mode is selected.

Margin for Negative Absorbance

The detector adjusts its gain during *balance* such that the baseline may drift slightly negative (about -100 mAU). In some special case, for example, when gradient with absorbing solvents are used, the baseline may drift to more negative values.

Only for such cases, increase the margin for negative absorbance to avoid overflow of the analog-to-digital converter.

Optimizing Selectivity

Quantifying Coeluting Peaks by Peak Suppression

In chromatography, two compounds may often elute together. A conventional dual-signal detector can only detect and quantify both compounds independently from each other if their spectra do not overlap. However, in most cases this is highly unlikely.

With a dual-channel detector based on diode-array technology, quantifying two compounds is possible even when both compounds absorb over the whole wavelength range. The procedure is called peak suppression or signal subtraction. As an example, the analysis of hydrochlorothiazide in the presence of caffeine is described. If hydrochlorothiazide is analyzed in biological samples, there is always a risk that caffeine is present which might interfere chromatographically with hydrochlorothiazide. As the spectra in [Figure 50](#) on page 138 shows, hydrochlorothiazide is best detected at 222 nm, where caffeine also shows significant absorbance. It would therefore be impossible, with a conventional variable wavelength detector, to detect hydrochlorothiazide quantitatively when caffeine is present.




Figure 50 Wavelength Selection for Peak Suppression

With a UV-visible detector based on a diode array and the correct choice of a reference wavelength setting, quantitative detection is possible. To suppress caffeine, the reference wavelength must be set to 282 nm. At this wavelength, caffeine shows exactly the same absorbance as at 222 nm. When the absorbance values are subtracted from each other, any indication of the presence of caffeine is eliminated. In the same way, hydrochlorothiazide can be suppressed if caffeine is to be quantified. In this case the wavelength is set to 204 nm and the reference wavelength to 260 nm. [Figure 51](#) on page 139 shows the chromatographic results of the peak suppression technique.

The trade-off for this procedure is a loss in sensitivity. The sample signal decreases by the absorbance at the reference wavelength relative to the signal wavelength. Sensitivity may be decreased by as much as 10–30 %.




Figure 51 Peak Suppression Using Reference Wavelength

Ratio Qualifiers for Selective Detection of Compound Classes

Ratio qualifiers can be used where, in a complex sample, only one particular class needs to be analyzed — a parent drug and its metabolites in a biological sample, for example. Another example is the selective analysis of derivatives after pre- or post-column derivatization. Specifying a signal ratio that is typical for the sample class is one way of selectively plotting only those peaks that are of interest. The signal output remains at zero so long as the ratio is out of the user-specified ratio range. When the ratio falls within the range, the signal output corresponds to the normal absorbance, giving single, clear peaks on a flat baseline. An example is shown in [Figure 52](#) on page 140 and [Figure 53](#) on page 141.




Figure 52 Wavelength Selection for Ratio Qualifiers




Figure 53 Selectivity by Ratio Qualifiers

In a four-component mixture, only biphenyl was recorded. The other three peaks were suppressed because they did not meet the ratio-qualifier criterion and therefore the output was set to zero. The characteristic wavelengths 249 nm (λ_1) and 224 nm (λ_2) were found from the spectra shown in [Figure 52](#) on page 140. The ratio range was set at 2 – 2.4 ($2.2 \pm 10\%$). Only when the ratio between 249 and 224 nm was within this range, is the signal plotted. Of all four peaks, only the third fulfilled the criterion ([Figure 53](#) on page 141). The others were not plotted.

Optimizing the Detector Regarding to the System

Delay Volume and Extra-Column Volume

The *delay volume* is defined as the system volume between the point of mixing in the pump and the top of the column.

The *extra-column volume* is defined as the volume between the injection point and the detection point, excluding the volume in the column.

Extra-Column Volume

Extra-column volume is a source of peak dispersion that will reduce the resolution of the separation and so should be minimized. Smaller diameter columns require proportionally smaller extra-column volumes to keep peak dispersion at a minimum.

In a liquid chromatograph the extra-column volume will depend on the connection tubing between the autosampler, column and detector; and on the volume of the flow cell in the detector. The extra-column volume is minimized with the Agilent 1290 Infinity/Agilent 1260 Infinity LC System due to the narrow-bore (0.12 mm i.d.) tubing, the low-volume heat exchangers in the column compartment and the Max-Light cartridge cell in the detector.

How to Configure the Optimum Delay Volume

To maintain resolution in the Agilent 1290 Infinity/Agilent 1260 Infinity Diode-array Detector the Max-Light cartridge cell has a low dispersion volume (σ volume 1.0 μl) and no further volume optimization is required. In situations where the alternative Agilent Max-Light high sensitivity cell is used to get higher sensitivity the cell volume is optimized for the use with 3 mm and 4.6 mm inner diameter columns.

How to Achieve Higher Sensitivity

The detector has a number of parameters that are used to optimize performance. The following sections describe how the detector parameters affect performance characteristics:

- Flow cell affects sensitivity,
- Wavelength and bandwidth affect sensitivity, selectivity and linearity,
- Slit Width affects sensitivity, spectral resolution and linearity,
- Peak Width affects sensitivity and resolution.

Flow Cell

The Max-Light cartridge flow cell has a standard 10 mm path length and is optimized for minimal volume and dispersion (σ volume 1.0 μl). It has high light transmission minimizing noise to reduce noise due to the optofluidic waveguide. It is suitable for use with a wide range of analytical columns from short narrow-bore columns to long standard diameter (4.6 mm) columns. Generally the peak dispersion volume (calculated from peak width x flow rate) should be greater than about 2 μl for this cell (for example 0.02 min x 200 $\mu\text{l}/\text{min}$ = 4 μl).

The Max-Light high sensitivity cell has a path length of 60 mm and this will give between three and five times increase in signal-to-noise values depending on the application conditions. The dispersion volume is fractionally increased compared to the standard cell.

Wavelength and Bandwidth

The detector measures absorbance simultaneously at wavelengths from 190 nm to 640 nm using diode-array detection. A UV-lamp provides good sensitivity over the whole wavelength range. The diode-array detector (DAD) can simultaneously compute and send to the data system up to eight chromatographic signals and the full-range spectra at every time point. A multi-wavelength detector (MWD) processes only the chromatographic signals. A UV chromatogram or signal is a plot of absorbance data versus time and is defined by its wavelength and bandwidth.

- The wavelength indicates the center of the detection band.
- The bandwidth defines the wavelength range over which the absorbance values are averaged to give the result at each time point.

For example, a signal at wavelength 250 nm with a bandwidth of 16 nm will be an average of the absorbance data from 242 nm to 258 nm. Additionally, a reference wavelength and reference bandwidth can be defined for each signal. The average absorbance from the reference bandwidth centered on the reference wavelength will be subtracted from its equivalent value at the signal wavelength to produce the output chromatogram.

The signal wavelength and bandwidth can be chosen so that they are optimized for:

- Broad band universal detection
- Narrow band selective detection
- Sensitivity for a specific analyte.

Broad band or universal detection works by having a wide bandwidth to detect any species with absorbance in that range. For example, to detect all absorbing molecules between 200 nm and 300 nm set a signal at 250 nm with a bandwidth of 100 nm. The disadvantage is that sensitivity will not be optimal for any one of those molecules. Narrow band or selective detection is used most often. The UV spectrum for a particular molecule is examined and an appropriate absorbance maximum is selected. If possible, the range where solvents absorb strongly should be avoided (below 220 nm for methanol, below 210 nm for acetonitrile). For example, in [Figure 54](#) on page 146, anisic acid has a suitable absorbance maximum at 252 nm. A narrow bandwidth of 4 nm to 12 nm generally gives good sensitivity and is specific for absorbance in a narrow range.

The narrow band can be optimized for sensitivity for a specific molecule. As the bandwidth is increased the signal is reduced but so is the noise and there will be an optimum for best S/N. As an approximate guide, this optimum is often close to the natural bandwidth at half-height of the absorption band in the UV spectrum. In the anisic acid example this is 30 nm.

The analytical wavelength is usually set at a wavelength maximum to increase sensitivity to that molecule. The detector is linear up to 2 AU and beyond for many applications. This offers a wide linear range for concentration. For high concentration analysis the concentration linear range can be extended by setting the wavelength to one with a lower absorbance such as a wavelength minimum or by taking a wider bandwidth which usually includes lower absorbance values. The use of wavelength maxima and minima for quantitation dates back to conventional UV detectors which because of mechanical tolerances in moving gratings needed to avoid steeply sloping parts of the spectrum. Diode-array based detectors do not have this limitation but for reasons of convention maxima and minima are chosen in preference to other parts of the spectrum.

The reference bandwidth is normally set on a region of the UV spectrum in which the analyte has no absorbance. This is shown in the spectrum for anisic acid in [Figure 54](#) on page 146. This spectrum is typical of many small molecules containing a UV chromophore. For best results the reference has been set so that it is a wide band as close to the signal wavelength as possible but on a zero absorbance region. Reference bandwidths of 60 nm to 100 nm are commonly used. The default reference is 360 nm with a bandwidth of 100 nm. A wide bandwidth is used because this reduces the noise in the reference signal (from statistical theory, the error, i.e. noise in this case, is reduced by the square root of the number of determinations). It is important that the reference bandwidth does not extend to a part of the spectrum that has some absorbance as this would then reduce the resulting signal and sensitivity would be reduced. The use of a reference wavelength can help to reduce drift or wander in the chromatogram caused by refractive index changes due to room temperature fluctuation or gradient operation. The effect of a reference signal can be easily tested by setting two otherwise identical signals, one with and one without a reference signal. If there is no part of the spectrum with zero absorbance then it will be better to have the reference signal turned off.

6 Optimizing the Detector

Optimizing the Detector Regarding to the System




Figure 54 Spectrum of Anisic Acid

Peak Width, Response Time and Data Collection Rate

The peak width setting, response time and data rate in the detector are all linked. The available settings are shown in [Table 21](#) on page 130. It is important to set this correctly for optimum sensitivity and to preserve the resolution achieved in the separation.

The detector internally acquires data points faster than is needed for a chromatogram and processes them to produce the signal seen by the data system. Part of the processing reduces the data to an appropriate data rate which allows the chromatographic peaks to be accurately drawn. As with most analytical determinations groups of readings are effectively averaged to reduce error in the result. The detector bunches raw data points and produces the output signal data at the required data collection rate by an electronic filtering process. If the resulting data rate is too slow (over filtering) the peak heights will be reduced and the resolution between them reduced; too fast and the data is noisier than it need be to accurately profile narrow peaks.

The *peak width* setting in the detector allows the user to correctly set these parameters without needing any knowledge other than sight of the chromatogram integration results to see how wide the peaks are. The peak width setting should be set for the narrowest peak width observed in the chromatogram. If it is set too wide it will make the peaks appear lower in height and wider (and potentially less resolved) and if it is set too narrow it will increase the baseline noise unnecessarily. Essentially the software uses this value to set the *data collection rate* such that it collects enough data points over the narrowest peaks and it is aiming for 15 to 25 points across a peak. The 1290 Infinity DAD can collect at a maximum 160 Hz if required which would allow enough data points to be collected over a peak that is only 0.1 s wide. The *response time* setting is another way of indicating how this filtering is set. It is measured in seconds and is about one-third of the peak width value (which is measured in minutes). It effectively shows how quickly the plotted signal responds to a step change in the input signal.

NOTE

The full spectra is not available under all conditions.

Based on the data points, the scan data rate is reduced, see [Table 21](#) on page 130.

6 Optimizing the Detector

Optimizing the Detector Regarding to the System

Table 22 Peak Width — Response Time — Data Rate

Peak width at half height [min] ¹	Response [s]	Signal data rate [Hz]	Scan data rate [HZ] ≤126 pts/scan	Scan data rate [HZ] ≤251 pts/scan	Scan data rate [HZ] ≤501 pts/scan	Scan data rate [HZ] >501 pts/scan
< 0.0016	0.016	160 ²	160 ²	80	40	20
> 0.0016	0.03	160 ²	160 ²	80	40	20
> 0.003	0.062	80	80	80	80	40
> 0.006	0.12	40	40	40	40	40
> 0.012	0.25	20	20	20	20	20
> 0.025	0.5	10	10	10	10	10
> 0.05	1.0	5	5	5	5	5
> 0.10	2.0	2.5	2.5	2.5	2.5	2.5
> 0.20	4.0	1.25	1.25	1.25	1.25	1.25
> 0.40	8.0	0.625	0.62	0.625	0.625	0.625
> 0.85	16.0	0.3125	0.31	0.3125	0.3125	0.3125

¹ Values in the User Interface may be rounded.

² G4212A only

NOTE

The maximum spectra scan rate depends on the data points per scan, see [Table 21](#) on page 130. Running at 160 Hz, the spectra scan data rate is reduced automatically if the spectra scan data rate is more than 251 points/scan.

Warm up of the Detector

Give the optical unit enough time to warm-up and stabilize (> 60 minutes). The detector is temperature controlled. After turn-on of the detector, it goes through a cycle of different states:

- 0 to 0.5 minutes the heater control is OFF and the heater element runs at 0 % duty cycle.
- 0.5 to 1 minutes the heater control is OFF and the heater element runs at 66% duty cycle. This first minute is used as self-test of the heater functionality.
- 1 to 30 minutes the heater control is OFF and the heater element runs at 40% duty cycle.
- After 30 minutes the heater control is ON and is working with optimized parameters to get the optical unit into the optimal temperature window stabilized.

This cycle starts

- when the detector is turned off/on
- when the lamp is turned off/on

to ensure that the temperature control operates in a defined control range.

NOTE

The times to stabilize the baseline may vary from instrument to instrument and depends on the environment. The example below was done under stable environmental conditions.

The figures below show the first two hours of a detector warm-up phase. The lamp was turned on immediately after turn on of the detector.

6 Optimizing the Detector

Warm up of the Detector




Figure 55 Detector Warm-up – 1st hour




Figure 56 Detector Warm-up – 2nd hour



7 Troubleshooting and Diagnostics

Overview of the Module's Indicators and Test Functions [152](#)

Status Indicators [153](#)

Power Supply Indicator [153](#)

Module Status Indicator [154](#)

Available Tests vs User Interfaces [155](#)

Agilent Lab Advisor Software [156](#)

Overview about the troubleshooting and diagnostic features.



Overview of the Module's Indicators and Test Functions

Status Indicators

The module is provided with two status indicators which indicate the operational state (prerun, run, and error states) of the module. The status indicators provide a quick visual check of the operation of the module.

Error Messages

In the event of an electronic, mechanical or hydraulic failure, the module generates an error message in the user interface. For each message, a short description of the failure, a list of probable causes of the problem, and a list of suggested actions to fix the problem are provided (see chapter Error Information).

Test Functions

A series of test functions are available for troubleshooting and operational verification after exchanging internal components (see Tests and Calibrations).

Diagnostic Signals

The module has several signals (internal temperatures, voltages and currents of lamps) that can be used for diagnosing baseline problems; see Diagnostic Signals.

Status Indicators

Two status indicators are located on the front of the module. The lower left indicates the power supply status, the upper right indicates the module status.




Figure 57 Location of Status Indicators

Power Supply Indicator

The power supply indicator is integrated into the main power switch. When the indicator is illuminated (*green*) the power is *ON*.

Module Status Indicator

The module status indicator indicates one of six possible module conditions:

- When the status indicator is *OFF* (and power switch light is on), the module is in a *prerun* condition, and is ready to begin an analysis.
- A *green* status indicator, indicates the module is performing an analysis (*run* mode).
- A *yellow* indicator indicates a *not-ready* condition. The module is in a not-ready state when it is waiting for a specific condition to be reached or completed (for example, immediately after changing a set point), or while a self-test procedure is running.
- An *error* condition is indicated when the status indicator is *red*. An error condition indicates the module has detected an internal problem which affects correct operation of the module. Usually, an error condition requires attention (e.g. leak, defective internal components). An error condition always interrupts the analysis.
- A *red-blinking* (modules with on-board LAN) or *yellow-blinking* (modules without on-board LAN) indicator indicates that the module is in resident mode (e.g. during update of main firmware).
- A *fast red-blinking* (modules with on-board LAN) or *fast yellow-blinking* (modules without on-board LAN) indicator indicates that the module is in boot loader mode (e.g. during update of main firmware). In such a case try to re-boot the module or try a cold-start.

Available Tests vs User Interfaces

- Depending on the user interface, the available tests and the screens/reports may vary (see Chapter *"Test Functions and Calibrations"*).
- Preferred tool should be the Agilent Diagnostic Software, see [“Agilent Lab Advisor Software”](#) on page 156.
- The Agilent ChemStation B.04.02 and above do not include any maintenance/test functions.
- Screenshots used within these procedures are based on the Agilent Lab Advisor Software.


Agilent Lab Advisor Software

The Agilent Lab Advisor software is a standalone product that can be used with or without data system. Agilent Lab Advisor software helps to manage the lab for high quality chromatographic results and can monitor in real time a single Agilent LC or all the Agilent GCs and LCs configured on the lab intranet.

Agilent Lab Advisor software provides diagnostic capabilities for all Agilent 1200 Infinity Series modules. This includes diagnostic capabilities, calibration procedures and maintenance routines for all the maintenance routines.

The Agilent Lab Advisor software also allows users to monitor the status of their LC instruments. The Early Maintenance Feedback (EMF) feature helps to carry out preventive maintenance. In addition, users can generate a status report for each individual LC instrument. The tests and diagnostic features as provided by the Agilent Lab Advisor software may differ from the descriptions in this manual. For details refer to the Agilent Lab Advisor software help files.

This manual provides lists with the names of Error Messages, Not Ready messages, and other common issues.



8 Error Information

What Are Error Messages	158
General Error Messages	159
Timeout	159
Shut-Down	160
Remote Timeout	160
Synchronization Lost	161
Leak Sensor Short	161
Leak Sensor Open	162
Compensation Sensor Open	162
Compensation Sensor Short	163
Fan Failed	163
Leak	164
Open Cover	164
Cover Violation	165
Detector Error Messages	166
Diode Current Leakage	166
UV Lamp Current	166
UV Lamp Voltage	167
UV Ignition Failed	167
UV Heater Current	168
Calibration Values Invalid	168
Wavelength Recalibration Lost	169
Illegal Value From Temperature Sensor	169
Illegal Value From Air Inlet Temperature Sensor	170
Heater Failed	170
Heater Power At Limit	171

This chapter describes the meaning of error messages, and provides information on probable causes and suggested actions how to recover from error conditions.



What Are Error Messages

Error messages are displayed in the user interface when an electronic, mechanical, or hydraulic (flow path) failure occurs which requires attention before the analysis can be continued (for example, repair, or exchange of consumables is necessary). In the event of such a failure, the red status indicator at the front of the module is switched on, and an entry is written into the module logbook.

General Error Messages

General error messages are generic to all Agilent series HPLC modules and may show up on other modules as well.

Timeout

The timeout threshold was exceeded.

Probable cause

- 1 The analysis was completed successfully, and the timeout function switched off the module as requested.
- 2 A not-ready condition was present during a sequence or multiple-injection run for a period longer than the timeout threshold.

Suggested actions

Check the logbook for the occurrence and source of a not-ready condition. Restart the analysis where required.

Check the logbook for the occurrence and source of a not-ready condition. Restart the analysis where required.

Shut-Down

An external instrument has generated a shut-down signal on the remote line.

The module continually monitors the remote input connectors for status signals. A LOW signal input on pin 4 of the remote connector generates the error message.

Probable cause	Suggested actions
1 Leak detected in another module with a CAN connection to the system.	Fix the leak in the external instrument before restarting the module.
2 Leak detected in an external instrument with a remote connection to the system.	Fix the leak in the external instrument before restarting the module.
3 Shut-down in an external instrument with a remote connection to the system.	Check external instruments for a shut-down condition.

Remote Timeout

A not-ready condition is still present on the remote input. When an analysis is started, the system expects all not-ready conditions (for example, a not-ready condition during detector balance) to switch to run conditions within one minute of starting the analysis. If a not-ready condition is still present on the remote line after one minute the error message is generated.

Probable cause	Suggested actions
1 Not-ready condition in one of the instruments connected to the remote line.	Ensure the instrument showing the not-ready condition is installed correctly, and is set up correctly for analysis.
2 Defective remote cable.	Exchange the remote cable.
3 Defective components in the instrument showing the not-ready condition.	Check the instrument for defects (refer to the instrument's documentation).

Synchronization Lost

During an analysis, the internal synchronization or communication between one or more of the modules in the system has failed.

The system processors continually monitor the system configuration. If one or more of the modules is no longer recognized as being connected to the system, the error message is generated.

Probable cause

- 1 CAN cable disconnected.
- 2 Defective CAN cable.
- 3 Defective main board in another module.

Suggested actions

- Ensure all the CAN cables are connected correctly.
 - Ensure all CAN cables are installed correctly.
- Exchange the CAN cable.
- Switch off the system. Restart the system, and determine which module or modules are not recognized by the system.

Leak Sensor Short

The leak sensor in the module has failed (short circuit).

The current through the leak sensor is dependent on temperature. A leak is detected when solvent cools the leak sensor, causing the leak-sensor current to change within defined limits. If the current increases above the upper limit, the error message is generated.

Probable cause

- 1 Defective flow sensor.
- 2 Leak sensor incorrectly routed, being pinched by a metal component.

Suggested actions

- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

Leak Sensor Open

The leak sensor in the module has failed (open circuit).

The current through the leak sensor is dependent on temperature. A leak is detected when solvent cools the leak sensor, causing the leak-sensor current to change within defined limits. If the current falls outside the lower limit, the error message is generated.

Probable cause

- 1** Leak sensor not connected to the main board.
- 2** Defective leak sensor.
- 3** Leak sensor incorrectly routed, being pinched by a metal component.

Suggested actions

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Compensation Sensor Open

The ambient-compensation sensor (NTC) on the main board in the module has failed (open circuit).

The resistance across the temperature compensation sensor (NTC) on the main board is dependent on ambient temperature. The change in resistance is used by the leak circuit to compensate for ambient temperature changes. If the resistance across the sensor increases above the upper limit, the error message is generated.

Probable cause

- 1** Defective main board.

Suggested actions

Please contact your Agilent service representative.

Compensation Sensor Short

The ambient-compensation sensor (NTC) on the main board in the module has failed (short circuit).

The resistance across the temperature compensation sensor (NTC) on the main board is dependent on ambient temperature. The change in resistance is used by the leak circuit to compensate for ambient temperature changes. If the resistance across the sensor falls below the lower limit, the error message is generated.

Probable cause

- 1 Defective main board.

Suggested actions

Please contact your Agilent service representative.

Fan Failed

The cooling fan in the module has failed.

The hall sensor on the fan shaft is used by the main board to monitor the fan speed. If the fan speed falls below a certain limit for a certain length of time, the error message is generated.

Probable cause

- 1 Fan cable disconnected.
- 2 Defective fan.
- 3 Defective main board.

Suggested actions

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Leak

A leak was detected in the module.

The signals from the two temperature sensors (leak sensor and board-mounted temperature-compensation sensor) are used by the leak algorithm to determine whether a leak is present. When a leak occurs, the leak sensor is cooled by the solvent. This changes the resistance of the leak sensor which is sensed by the leak-sensor circuit on the main board.

Probable cause

- 1** Loose fittings.
- 2** Broken capillary.

Suggested actions

Ensure all fittings are tight.

Exchange defective capillaries.

Open Cover

The top foam has been removed.

Probable cause

- 1** Foam not activating the sensor.
- 2** Dirty or defective sensor.

Suggested actions

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Cover Violation

The top foam has been removed.

The sensor on the main board detects when the top foam is in place. If the foam is removed while the lamp is on (or if an attempt is made to switch on for example the lamp with the foam removed), the lamp is switched off, and the error message is generated.

Probable cause

- 1 The top foam was removed during operation.
- 2 Foam not activating the sensor.

Suggested actions

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Detector Error Messages

These errors are detector specific.

Diode Current Leakage

When the detector is switched on, the processor checks the leakage current of each of the optical diodes. If the leakage current exceeds the upper limit, the error message is generated.

Probable cause

- 1 Defective PDA/optical unit.
- 2 Defective connector or cable.

Suggested actions

Please contact your Agilent service representative.

Please contact your Agilent service representative.

UV Lamp Current

The UV lamp current is missing.

The processor continually monitors the anode current drawn by the lamp during operation. If the anode current falls below the lower current limit, the error message is generated.

Probable cause

- 1 Lamp disconnected.
- 2 Defective UV lamp or non-Agilent lamp.
- 3 Defective detector main board.
- 4 Defective power supply.

Suggested actions

Ensure the UV lamp connector is seated firmly.

Exchange the UV lamp.

Please contact your Agilent service representative.

Please contact your Agilent service representative.

UV Lamp Voltage

The UV lamp anode voltage is missing.

The processor continually monitors the anode voltage across the lamp during operation. If the anode voltage falls below the lower limit, the error message is generated.

Probable cause

- 1 Defective UV lamp or non-Agilent lamp.
- 2 Defective detector main board.
- 3 Defective power supply.

Suggested actions

- Exchange the UV lamp.
- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

UV Ignition Failed

The UV lamp failed to ignite.

The processor monitors the UV lamp current during the ignition cycle. If the lamp current does not rise above the lower limit within 2 – 5 seconds, the error message is generated.

Probable cause

- 1 Lamp too hot. Hot gas discharge lamps may not ignite as easily as cold lamps.
- 2 Lamp disconnected.
- 3 Defective UV lamp or non-Agilent lamp.
- 4 Defective detector main board.
- 5 Defective power supply.

Suggested actions

- Switch off the lamp and allow it to cool down for at least 15 minutes.
- Ensure the lamp is connected.
- Exchange the UV lamp.
- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

UV Heater Current

The UV lamp heater current is missing.

During UV lamp ignition, the processor monitors the heater current. If the current does not rise above the lower limit within one second, the error message is generated.

Probable cause	Suggested actions
1 Lamp disconnected.	Ensure the UV lamp is connected.
2 Ignition started without the top foam in place.	Please contact your Agilent service representative.
3 Defective UV lamp or non-Agilent lamp.	Exchange the UV lamp.
4 Defective detector main board.	Please contact your Agilent service representative.
5 Defective power supply.	Please contact your Agilent service representative.

Calibration Values Invalid

The calibration values read from the spectrometer ROM are invalid.

After recalibration, the calibration values are stored in ROM. The processor periodically checks if the calibration data are valid. If the data are invalid or cannot be read from the spectrometer ROM, the error message is generated.

Probable cause	Suggested actions
1 Defective connector or cable.	Please contact your Agilent service representative.
2 Defective PDA/optical unit.	Please contact your Agilent service representative.

Wavelength Recalibration Lost

The calibration information needed for your detector to operate correctly has been lost.

During calibration of the detector the calibration values are stored in ROM. If no data is available in the spectrometer ROM, the error message is generated.

Probable cause

- 1 The detector is new.
- 2 The detector has been repaired.

Suggested actions

Recalibrate the detector.

Please contact your Agilent service representative.

Illegal Value From Temperature Sensor

This temperature sensor delivered a value outside the allowed range. The parameter of this event equals the measured temperature in 1/100 centigrade. As a result the temperature control is switched off.

Probable cause

- 1 Dirty or defective sensor.
- 2 Detector is exposed to illegal ambient conditions.

Suggested actions

Please contact your Agilent service representative. Links:

Verify that the ambient conditions are within the allowed range.

Illegal Value From Air Inlet Temperature Sensor

This temperature sensor (located on the detector main board) delivered a value outside the allowed range. The parameter of this event equals the measured temperature in 1/100 centigrade. As a result the temperature control is switched off.

Probable cause

- 1 The temperature sensor is defect.
- 2 Detector is exposed to illegal ambient conditions.

Suggested actions

Please contact your Agilent service representative.

Verify that the ambient conditions are within the allowed range.

Heater Failed

Every time the deuterium lamp or the tungsten lamp is switched on or off a heater self-test is performed. If the test fails an error event is created. As a result the temperature control is switched off.

Probable cause

- 1 Defective connector or cable.
- 2 Defective heater.

Suggested actions

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Heater Power At Limit

The available power of the heater reached either the upper or lower limit. This event is sent only once per run. The parameter determines which limit has been hit:

0 means upper power limit hit (excessive ambient temperature drop).

1 means lower power limit hit (excessive ambient temperature increase).

Probable cause


- 1 Ambient conditions have changed too much during the run, so that optimum results may not be guaranteed.

Suggested actions

- Verify that the reproducibility of your results is not affected.
- Expose the detector to more stable ambient conditions.

8 Error Information

Detector Error Messages



9 Test Functions and Calibration

Introduction	174
Use of Max-Light Cartridge Test Cell	175
Conditions of Detector	176
Failing a Test	177
Self-Test	178
Intensity Test	180
Test Failed	182
Cell Test	183
Test Failed (low ratio value)	184
Quick Noise Test	185
Test Failed	187
ASTM Drift and Noise Test	188
Test Failed	190
Slit Test (G4212A)	191
Test Failed	192
Wavelength Verification Test	193
Wavelength Calibration	195
Wavelength Recalibration Fails	197
D/A Converter (DAC) Test	198
Test Failed	199
Dark Current Test	200
Test Failed	201

This chapter describes the tests for the module.



Introduction

All tests are described based on the Agilent Lab Advisor Software B.01.03. Other user interfaces may not provide any test or just a few.

Table 23 Interfaces and available test functions

Interface	Comment	Available Function
Agilent Instrument Utilities	Maintenance tests are available	<ul style="list-style-type: none"> • Intensity • Cell • WL Calibration
Agilent Lab Advisor	All tests are available	<ul style="list-style-type: none"> • Self-Test • Intensity • Quick Noise • ASTM Drift and Noise • Cell • Dark Current • D/A Converter • Slit (G4212A only) • WL Verification • WL Calibration • Test Chromatogram (Tools) • Spectra Scan (Tools) • Module Infos (Tools) • Diagnostic (Tools)
Agilent ChemStation	No tests available Adding of temperature/lamp signals to chromatographic signals possible	<ul style="list-style-type: none"> • Temperature main board • Temperature optical unit • Lamp anode voltage
Agilent Instant Pilot	Some tests are available	<ul style="list-style-type: none"> • Intensity • WL Calibration • Cell

For details on the use of the interface refer to the interface documentation.

Use of Max-Light Cartridge Test Cell

The Test Cell Cartridge provided with every detector is recommended to be used for several tests instead of the used Max-Light Cartridge Cell (10 mm $\sigma = 1 \mu\text{l}$ or 60 mm $\sigma = 4 \mu\text{l}$) because it allows running the test(s) without any influence of the rest of the system (degasser, pump, sampler and others).

The results of the test cell are comparable with the Max-Light Cartridge Cell 10 mm $\sigma = 1 \mu\text{l}$ filled with water, e.g. Intensity Profile. Only the Absorbance value is higher on the Max-Light Cartridge Cell.

If the profile of the Max-Light Cartridge Cell differs in the low UV range, then absorbing solvents are in the cell and should be flushed out. See also “[Cleaning of Max-Light Cartridge Cell](#)” on page 218.

NOTE

When using the Max-Light Cartridge Cell for tests/calibrations, it should be run at 0.5 ml/min constant flow with water. This assures that the light path is always flushed.

Below table gives an idea on the signal height variation of the Max-Light Cartridge Cells compared to Max-Light Cartridge Test Cell.

Table 24 Max-Light Cartridge Cells compared to Max-Light Cartridge Test Cell

Part Number	Description	Signal Height (typical)
G4212-60011	Max-Light Cartridge Test Cell	100 %
G4212-60008	Max-Light Cartridge Cell 10 mm $V(\sigma) = 1 \mu\text{l}$	~ 100 %
G4212-60007	Max-Light Cartridge Cell 60 mm $V(\sigma) = 4 \mu\text{l}$	~ 100 %

Conditions of Detector

The test usually should be performed with a detector turned on for at least one hour, so that the temperature regulation of the optical unit is working (not active during the first 30 minutes after turn on). If the detector is on, tests can be performed usually 10 minutes after the UV-lamp has been turned on.

Failing a Test

If a test fails with the Max-Light Cartridge Cell repeat the test with the Max-Light Cartridge Test Cell and compare. If the test fails also, then start with proposed actions mentioned in the details of the tests.

Self-Test

The self-test runs a series of individual tests (described on the next pages), and evaluates the results automatically. The following tests are run:

- Slit Test
- Dark Current Test
- Intensity Test
- Wavelength Verification Test
- ASTM Noise Test, a simplified version of the ASTM Drift and Noise Test (without testing the Drift)

When For complete detector check.

Parts required	#	Description
	1	Max-Light Cartridge Cell (filled with water) or
	1	Max-Light Cartridge Test Cell

- Preparations**
- Lamp must be on for at least 10 minutes.
 - For noise test a longer warm-up time may be required (> 2 hours).
 - When using a Max-Light Cartridge Cell a flow rate of 0.5 ml/min with water is required.

- 1 Run the **Self-Test** with the recommended user interface (for further information see Online-Help of user interface).

Test Name	Self Test	Description	The test performs a self test.
Module	G4212A-PR00100015		
Status	Passed		
Start Time	7/9/2009 2:21:51 PM		
Stop Time	7/9/2009 2:43:51 PM		

Test Procedure	Result																																														
✓ 1. Check Prerequisites...	<table> <tr> <th>Name</th><th>Value</th></tr> <tr> <td>Cell Product Number</td><td>G4212-60011</td></tr> <tr> <td>Cell Name</td><td>Max-Light Test Cell</td></tr> <tr> <td>Cell Type</td><td>10 mm/0 µl</td></tr> <tr> <td>Lamp Type</td><td>Automatic Mode</td></tr> <tr> <td>Slit Test Result</td><td>0.84</td></tr> <tr> <td>Dark Current Minimum</td><td>7699 Counts</td></tr> <tr> <td>Dark Current Maximum</td><td>7763 Counts</td></tr> <tr> <td>Lowest Intensity in Range 190 - 220 nm</td><td>30257 Counts</td></tr> <tr> <td>Lowest Intensity in Range 221 - 350 nm</td><td>35216 Counts</td></tr> <tr> <td>Lowest Intensity in Range 351 - 500 nm</td><td>8219 Counts</td></tr> <tr> <td>Lowest Intensity in Range 501 - 640 nm</td><td>2202 Counts</td></tr> <tr> <td>Highest Intensity in Range 190 - 350 nm</td><td>151632 Counts</td></tr> <tr> <td>Highest Intensity in Range 351 - 500 nm</td><td>38283 Counts</td></tr> <tr> <td>Highest Intensity in Range 501 - 640 nm</td><td>38703 Counts</td></tr> <tr> <td>D2 Alpha Line Deviation</td><td>0.000 nm</td></tr> <tr> <td>D2 Beta Line Deviation</td><td>0.000 nm</td></tr> <tr> <td>D2 Alpha Line</td><td>656.100 nm</td></tr> <tr> <td>D2 Beta Line</td><td>486.000 nm</td></tr> <tr> <td>Spectral Flatness</td><td>0.0001 AU</td></tr> <tr> <td>Accumulated UV Lamp Burn Time</td><td>84.00 h</td></tr> <tr> <td>UV Lamp On-Time</td><td>4.84 h</td></tr> <tr> <td>Signal Noise value at 254 nm (UV)</td><td>0.007 mAU</td></tr> </table>	Name	Value	Cell Product Number	G4212-60011	Cell Name	Max-Light Test Cell	Cell Type	10 mm/0 µl	Lamp Type	Automatic Mode	Slit Test Result	0.84	Dark Current Minimum	7699 Counts	Dark Current Maximum	7763 Counts	Lowest Intensity in Range 190 - 220 nm	30257 Counts	Lowest Intensity in Range 221 - 350 nm	35216 Counts	Lowest Intensity in Range 351 - 500 nm	8219 Counts	Lowest Intensity in Range 501 - 640 nm	2202 Counts	Highest Intensity in Range 190 - 350 nm	151632 Counts	Highest Intensity in Range 351 - 500 nm	38283 Counts	Highest Intensity in Range 501 - 640 nm	38703 Counts	D2 Alpha Line Deviation	0.000 nm	D2 Beta Line Deviation	0.000 nm	D2 Alpha Line	656.100 nm	D2 Beta Line	486.000 nm	Spectral Flatness	0.0001 AU	Accumulated UV Lamp Burn Time	84.00 h	UV Lamp On-Time	4.84 h	Signal Noise value at 254 nm (UV)	0.007 mAU
Name	Value																																														
Cell Product Number	G4212-60011																																														
Cell Name	Max-Light Test Cell																																														
Cell Type	10 mm/0 µl																																														
Lamp Type	Automatic Mode																																														
Slit Test Result	0.84																																														
Dark Current Minimum	7699 Counts																																														
Dark Current Maximum	7763 Counts																																														
Lowest Intensity in Range 190 - 220 nm	30257 Counts																																														
Lowest Intensity in Range 221 - 350 nm	35216 Counts																																														
Lowest Intensity in Range 351 - 500 nm	8219 Counts																																														
Lowest Intensity in Range 501 - 640 nm	2202 Counts																																														
Highest Intensity in Range 190 - 350 nm	151632 Counts																																														
Highest Intensity in Range 351 - 500 nm	38283 Counts																																														
Highest Intensity in Range 501 - 640 nm	38703 Counts																																														
D2 Alpha Line Deviation	0.000 nm																																														
D2 Beta Line Deviation	0.000 nm																																														
D2 Alpha Line	656.100 nm																																														
D2 Beta Line	486.000 nm																																														
Spectral Flatness	0.0001 AU																																														
Accumulated UV Lamp Burn Time	84.00 h																																														
UV Lamp On-Time	4.84 h																																														
Signal Noise value at 254 nm (UV)	0.007 mAU																																														
✓ 2. Insert supported Cell or Test Cell																																															
✓ 3. Perform Slit Test...																																															
✓ 4. Perform Dark Current Test...																																															
✓ 5. Perform Intensity Test...																																															
✓ 6. Perform Wavelength Calibration Test...																																															
✓ 7. Perform Spectral Flatness Test...																																															
✓ 8. Perform ASTM Noise Test (20 min. at 254 nm)...																																															
✓ 9. Evaluate Data...																																															

Figure 58 Self-Test – Results

Intensity Test

The intensity test measures the intensity of the UV-lamp over the full wavelength range (190 - 640 nm). Four spectral ranges are used to evaluate the intensity spectrum. The test is used to determine the performance of the lamp and optics (see also “Cell Test” on page 183). When the test is started, the 1-nm slit is moved into the light path automatically (G4212A). To eliminate effects due to absorbing solvents, the test should be done with water in the Max-Light Cartridge Cell or with the Max-Light Cartridge Test Cell. The shape of the intensity spectrum is primarily dependent on the lamp, grating, and diode array characteristics. Therefore, intensity spectra will differ slightly between instruments.

When	In case of UV-lamp problem (drift, noise).						
Parts required	<table><tr><th>#</th><th>Description</th></tr><tr><td>1</td><td>Max-Light Cartridge Cell (filled with water) or</td></tr><tr><td>1</td><td>Max-Light Cartridge Test Cell</td></tr></table>	#	Description	1	Max-Light Cartridge Cell (filled with water) or	1	Max-Light Cartridge Test Cell
#	Description						
1	Max-Light Cartridge Cell (filled with water) or						
1	Max-Light Cartridge Test Cell						
Preparations	Lamp must be on for at least 10 minutes.						

- 1 Run the **Intensity-Test** with the recommended user interface (for further information see Online-Help of user interface).




Figure 59 Intensity Test – Results




Figure 60 Intensity Test – Signals

Test Failed

Intensity Test Evaluation

Probable cause

- 1** Absorbing solvent or air bubble in flow cell.
- 2** Incorrect calibration
- 3** Dirty or contaminated flow cell.
- 4** Dirty or contaminated optical components.
- 5** Old UV-lamp.
- 6** Defect optical unit.

Suggested actions

- Ensure the flow cell is filled with water, and free from air bubbles.
 - Repeat test with Max-Light Cartridge Test Cell and compare results.
- Recalibrate and repeat the test.
- Run the cell test. If the test fails, flush the flow cell. See also [“Cleaning of Max-Light Cartridge Cell”](#) on page 218.
- Please contact your Agilent service representative.
- Exchange the UV-lamp.
- If the test fails with Max-Light Cartridge Test Cell and new UV-lamp, please contact your Agilent service representative.

NOTE

If only one range fails and the application does not require this range, the lamp may not be changed.

Cell Test

The cell test measures the intensity of the UV-lamp over the full wavelength range (190 - 690 nm), once with the Max-Light Cartridge Cell installed, and once with the Max-Light Cartridge Test Cell. The resulting intensity ratio is a measure of the amount of light absorbed by the Max-Light Cartridge Test Cell. The test can be used to check for dirty or contaminated flow cell windows. When the test is started, the 1-nm slit is moved into the light path automatically (G4212A only). On the G4212B, the 4 nm fixed slit is used.

This test should be performed initially with a new detector/flow cell. The values should be kept for later reference/comparison.

When In case of low intensity or noise and drift problem.

Parts required

#	Description
1	Max-Light Cartridge Cell (filled with water) or
1	Max-Light Cartridge Test Cell

Preparations

- Lamp must be on for at least 10 minutes.
- When using a Max-Light Cartridge Cell a flow rate of 0.5 ml/min with water is required.

1 Run the **Cell-Test** with the recommended user interface (for further information see Online-Help of user interface).

Test Name	Cell Test	Description	The test compares the lamp intensity the Max-Light Cell and the Max-Light Test Cell. The intensity ratio is an indicator of the amount of light absorbed by the flow cell.																
Module	G4212A:PR00100018																		
Status	Passed																		
Start Time	7/14/2009 1:40:44 PM																		
Stop Time	7/14/2009 1:41:46 PM																		
<div><div></div></div>																			
Test Procedure		Result																	
<div><div></div><div></div><div></div><div></div><div></div><div></div></div> <div><div>1. Check Prerequisites...</div><div>2. Insert Test Cell.</div><div>3. Scan Intensity Spectrum...</div><div>4. Insert supported Cell.</div><div>5. Scan Intensity Spectrum...</div><div>6. Evaluate Data...</div></div>		<table><tr><th>Name</th><th>Value</th></tr><tr><td>Cell Product Number</td><td>G4212-60008</td></tr><tr><td>Cell Name</td><td>Max-Light Cell</td></tr><tr><td>Cell Type</td><td>10 mm/1 µl</td></tr><tr><td>Lamp Type</td><td>Automatic Mode</td></tr><tr><td>Intensity Integral with Test Cell</td><td>13,337,028</td></tr><tr><td>Intensity Integral with Flow Cell</td><td>15,661,215</td></tr><tr><td>Intensity Ratio</td><td>1.17</td></tr></table>		Name	Value	Cell Product Number	G4212-60008	Cell Name	Max-Light Cell	Cell Type	10 mm/1 µl	Lamp Type	Automatic Mode	Intensity Integral with Test Cell	13,337,028	Intensity Integral with Flow Cell	15,661,215	Intensity Ratio	1.17
Name	Value																		
Cell Product Number	G4212-60008																		
Cell Name	Max-Light Cell																		
Cell Type	10 mm/1 µl																		
Lamp Type	Automatic Mode																		
Intensity Integral with Test Cell	13,337,028																		
Intensity Integral with Flow Cell	15,661,215																		
Intensity Ratio	1.17																		

Figure 61 Cell Test – Results




Figure 62 Cell Test – Signals

Test Failed (low ratio value)

Cell Test Evaluation

Probable cause

- 1** Absorbing solvent or air bubble in flow cell.
- 2** Dirty or contaminated flow cell.

Suggested actions

Ensure the flow cell is filled with water, and free from air bubbles.

Clean the flow cell as described in [“Cleaning of Max-Light Cartridge Cell”](#) on page 218.

Quick Noise Test

The quick noise test measures the noise of the detector, with Max-Light Cartridge Cell or with Max-Light Cartridge Test Cell installed, in one minute intervals over a total of 5 minutes.

The noise of the detector is calculated by using the maximum amplitude for all random variations of the detector signal of frequencies greater than one cycle per hour. The noise is determined for 5 one minute intervals and is based on the accumulated peak-to-peak noise for the intervals. At least seven data points per cycles are used in the calculation. The cycles in the noise determination are not overlapping.

If the test is performed with the Max-Light Cartridge Test Cell, the test results are not influenced by solvent or pump effects.

When In case of noise and drift problem.

Parts required	<table border="0"> <tr> <td style="text-align: right;">#</td> <td>Description</td> </tr> <tr> <td style="text-align: right;">1</td> <td>Max-Light Cartridge Cell (filled with water) or</td> </tr> <tr> <td style="text-align: right;">1</td> <td>Max-Light Cartridge Test Cell</td> </tr> </table>	#	Description	1	Max-Light Cartridge Cell (filled with water) or	1	Max-Light Cartridge Test Cell
#	Description						
1	Max-Light Cartridge Cell (filled with water) or						
1	Max-Light Cartridge Test Cell						

Preparations	<ul style="list-style-type: none"> • Detector and UV-lamp must be on for at least 2 hours. • ASTM measurements based on specifications may require longer stabilization times. • When using a Max-Light Cartridge Cell a flow rate of 0.5 ml/min with water is required.
---------------------	---

9 Test Functions and Calibration
Quick Noise Test

- 1** Run the **Quick Noise Test** with the recommended user interface (for further information see Online-Help of user interface).

Test Name	Quick Noise Test	Description	The test performs a quick Noise Evaluation without reference.
Module	G4212A:PR00100015		
Status	Passed		
Start Time	7/9/2009 2:03:53 PM		
Stop Time	7/9/2009 2:09:10 PM		

Test Procedure		Result	
✓ 1.	Check Prerequisites...	Name	Value
✓ 2.	Insert supported Cell or Test Cell.	Cell Product Number	G4212-60011
✓ 3.	Measure Noise...	Cell Name	Max-Light Test Cell
✓ 4.	Evaluate Data...	Cell Type	10 mm/10 µl
		Lamp Type	Automatic Mode
		Accumulated UV Lamp Burn Time	83.68 h
		UV Lamp On-Time	4.51 h
		Signal Noise value at 254 nm (UV)	0.008 mAU

Figure 63 Quick Noise Test – Results




Figure 64 Quick Noise Test — Signals

Test Failed

Quick Noise Test Evaluation

Probable cause

- 1 Insufficient lamp warm-up time.
- 2 Absorbing solvent or air bubble in flow cell.
- 3 Dirty or contaminated flow cell.
- 4 Old UV-lamp.

Suggested actions

- Allow detector and UV-lamp turned on for at least 2 hours.
- Ensure the flow cell is filled with water, and free from air bubbles.
- Flush flow cell.
 - Clean the flow cell as described in [“Cleaning of Max-Light Cartridge Cell”](#) on page 218.
- Exchange the UV-lamp.

ASTM Drift and Noise Test

The ASTM noise test determines the detector noise over a period of 20 minutes. The test is done with installed Max-Light Cartridge Cell or Max-Light Cartridge Test Cell.

This test does also check for the drift. It is also part of the “Self Test” (without checking for the drift).

If the test is performed with the Max-Light Cartridge Test Cell, the test results are not influenced by solvent or pump effects.

When In case of noise and drift problem.

Parts required	#	Description
	1	Max-Light Cartridge Cell (filled with water) or
	1	Max-Light Cartridge Test Cell

Preparations

- Detector and UV-lamp must be on for at least 2 hours.
- ASTM measurements based on specifications may require longer stabilization times.
- When using a Max-Light Cartridge Cell a flow rate of 0.5 ml/min with water is required.

- 1 Run the **ASTM Drift and Noise Test** with the recommended user interface (for further information see Online-Help of user interface).




Figure 65 ASTM Drift and Noise Test – Results




Figure 66 ASTM Drift and Noise Test – Signals

Test Failed

ASTM Noise Test Evaluation

Probable cause

- 1** Insufficient lamp warm-up time.
- 2** Absorbing solvent or air bubble in flow cell.
- 3** Dirty or contaminated flow cell.
- 4** Old UV-lamp.
- 5** Environment not according to specifications.

Suggested actions

- Allow detector and UV-lamp turned on for at least 2 hours.
- Ensure the flow cell is filled with water, and free from air bubbles.
- Flush flow cell.
 - Clean the flow cell as described in [“Cleaning of Max-Light Cartridge Cell”](#) on page 218.
- Exchange the UV-lamp.
- Improve environment.

Slit Test (G4212A)

The slit test verifies correct operation of the micromechanical slit.

During the test, the slit is moved through all slit positions while the detector monitors the lamp intensity change. When the slit position is changed, the intensity drop (move to smaller slit) or intensity increase (move to larger slit) must be within a defined range.

If the intensity changes are outside the expected range, the test fails.

When

In case of problems.

Parts required

- | | |
|----------|---|
| # | Description |
| 1 | Max-Light Cartridge Cell (filled with water) or |
| 1 | Max-Light Cartridge Test Cell |

Preparations

- Lamp must be on for at least 10 minutes.
- When using a Max-Light Cartridge Cell a flow rate of 0.5 ml/min with water is required.

- 1 Run the **Slit Test** with the recommended user interface (for further information see Online-Help of user interface).


Test Name Slit Test Module G4212A:PR00100015 Status Passed Start Time 7/9/2009 2:19:17 PM Stop Time 7/9/2009 2:19:56 PM 	Description The test performs a slit test												
Test Procedure <ol style="list-style-type: none"> 1. Check Prerequisites... 2. Insert supported Cell or Test Cell. 3. Perform Slit Test... 4. Evaluate Data... 													
Result <table border="1"> <thead> <tr> <th>Name</th><th>Value</th></tr> </thead> <tbody> <tr> <td>Cell Product Number</td><td>G4212-60011</td></tr> <tr> <td>Cell Name</td><td>Max-Light Test Cell</td></tr> <tr> <td>Cell Type</td><td>10 mm/0 µl</td></tr> <tr> <td>Lamp Type</td><td>Automatic Mode</td></tr> <tr> <td>Slit Test Result</td><td>0.84</td></tr> </tbody> </table>		Name	Value	Cell Product Number	G4212-60011	Cell Name	Max-Light Test Cell	Cell Type	10 mm/0 µl	Lamp Type	Automatic Mode	Slit Test Result	0.84
Name	Value												
Cell Product Number	G4212-60011												
Cell Name	Max-Light Test Cell												
Cell Type	10 mm/0 µl												
Lamp Type	Automatic Mode												
Slit Test Result	0.84												

Figure 67 Slit Test – Results

9 Test Functions and Calibration

Slit Test (G4212A)




Figure 68 Slit Test – Signals

Test Failed

Slit Test Evaluation

Probable cause

- 1 Max-Light Cartridge Cell still installed.
- 2 Old lamp.
- 3 Defective slit assembly.
- 4 Defective detector main board.
- 5 Defective optical unit.

Suggested actions

Install Max-Light Cartridge Test Cell.

Run the "Intensity Test". Exchange the lamp if old or defective.

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Please contact your Agilent service representative.


Wavelength Verification Test

The detector uses the alpha (656.1 nm) and beta (486 nm) emission lines of the UV-lamp for wavelength calibration. The sharp emission lines enable accurate calibration. When verification is started, the 1-nm slit is moved into the light path automatically. The test is run with the Max-Light Cartridge Cell or with Max-Light Cartridge Test Cell installed.

If the test is performed with the Max-Light Cartridge Test Cell, the test results are not influenced by solvent or pump effects.

When	<p>The detector is calibrated at the factory, and under normal operating conditions should not require recalibration. However, it is advisable to recalibrate:</p> <ul style="list-style-type: none"> • after repair of components in the optical unit, • after exchange of the optical unit or main board, • after replacing the Max-Light Cartridge Cell or UV-lamp, • after significant environmental condition changes (temperature, humidity), • at a regular interval, at least once per year (for example, prior to an Operational Qualification/Performance Verification procedure), and • when chromatographic results indicate the detector may require recalibration. 						
Parts required	<table> <tr> <th>#</th><th>Description</th></tr> <tr> <td>1</td><td>Max-Light Cartridge Test Cell or</td></tr> <tr> <td>1</td><td>Max-Light Cartridge Cell</td></tr> </table>	#	Description	1	Max-Light Cartridge Test Cell or	1	Max-Light Cartridge Cell
#	Description						
1	Max-Light Cartridge Test Cell or						
1	Max-Light Cartridge Cell						
Preparations	<ul style="list-style-type: none"> • Lamp must be on for at least 10 minutes. • When using a Max-Light Cartridge Cell a flow rate of 0.5 ml/min with water is required. 						

- (for further information see Online-Help of user interface).

Test Name	Wavelength Verification Test	Description	The test performs a Wavelength Verification.
Module	G4212A:PR00100015		
Status	Passed		
Start Time	7/9/2009 1:50:40 PM		
Stop Time	7/9/2009 1:51:03 PM		
			

Test Procedure		Result	
		Name	Value
✓	1. Check Prerequisites...	Cell Product Number	G4212-60011
✓	2. Insert Test Cell	Cell Name	Max-Light Test Cell
✓	3. Wavelength Verification...	Cell Type	10 mm/0 µl
✓	4. Evaluate Data...	Lamp Type	Automatic Mode
		Previous Absolute D2 Alpha Line Deviation	0.000 nm
		Previous Absolute D2 Beta Line Deviation	0.000 nm
		D2 Alpha Line Deviation	-0.146 nm
		D2 Beta Line Deviation	-0.070 nm
		Absolute D2 Alpha Line Deviation	-0.146 nm
		Absolute D2 Beta Line Deviation	-0.070 nm
		D2 Alpha Line	655.954 nm
		D2 Beta Line	485.930 nm

Figure 69 Wavelength Verification – Results

Wavelength Calibration

The detector uses the alpha (656.1 nm) and beta (486 nm) emission lines of the deuterium lamp for wavelength calibration. The sharp emission lines enable more accurate calibration than is possible with holmium oxide. When recalibration is started, the 1 nm slit is moved into the light path automatically (G4212A). The gain is set to zero.

On completion of the scan, the alpha- and beta-line deviations (in nm) are displayed. These values indicate how far the detector calibration deviates from the actual positions of the alpha and beta emission lines. After calibration, the deviation is zero.

To eliminate effects due to absorbing solvents, install the Max-Light Cartridge Test Cell before starting the test.

When	<p>The detector is calibrated at the factory, and under normal operating conditions should not require recalibration. However, it is advisable to recalibrate:</p> <ul style="list-style-type: none"> • after maintenance (flow cell or UV-lamp), • after repair of components in the optical unit, • after exchange of the optical unit or main board, • after significant environmental condition changes (temperature, humidity), • at a regular interval, at least once per year (for example, prior to an Operational Qualification/Performance Verification procedure), and • when chromatographic results indicate the detector may require recalibration. 						
Parts required	<table> <tr> <th>#</th><th>Description</th></tr> <tr> <td>1</td><td>Max-Light Cartridge Test Cell or</td></tr> <tr> <td>1</td><td>Max-Light Cartridge Cell</td></tr> </table>	#	Description	1	Max-Light Cartridge Test Cell or	1	Max-Light Cartridge Cell
#	Description						
1	Max-Light Cartridge Test Cell or						
1	Max-Light Cartridge Cell						
Preparations	<ul style="list-style-type: none"> • Detector/lamp must be on for more than 1 hour. • When using a Max-Light Cartridge Cell a flow rate of 0.5 ml/min with water is required. 						

NOTE

If the detector is operated in a lab environment that differs at average from the final test environment (25 °C) then the detector should be recalibrated for this temperature.

NOTE

If the detector was repaired (opened covers), the wavelength calibration can be done 10 minutes after lamp on. A final wavelength calibration should be repeated after complete warm-up of the detector.

- 1 Run the **Wavelength Calibration** with the recommended user interface (for further information see Online-Help of user interface).


Test Name Wavelength Calibration Module G4212A:PR00100015 Status Passed Start Time 7/9/2009 1:59:26 PM Stop Time 7/9/2009 1:59:50 PM 	Description The wavelength calibration procedure enables you to check the calibration of the diode array in the detector. Calibration means adjusting the assignment of diodes to specific wavelengths, and is done using the two deuterium emission lines at 486.0 nm and 656.1 nm.																												
Test Procedure <ul style="list-style-type: none"> ✓ 1. Check Prerequisites... ✓ 2. Insert Test Cell. ✓ 3. Wavelength Verification... ✓ 4. Calibrate Detector... 	Result <table> <thead> <tr> <th>Name</th><th>Value</th></tr> </thead> <tbody> <tr> <td>Cell Product Number</td><td>G4212-60011</td></tr> <tr> <td>Cell Name</td><td>Max-Light Test Cell</td></tr> <tr> <td>Cell Type</td><td>10 mm/0 µl</td></tr> <tr> <td>Lamp Type</td><td>Automatic Mode</td></tr> <tr> <td>Previous Absolute D2 Alpha Line Deviation</td><td>0.000 nm</td></tr> <tr> <td>Previous Absolute D2 Beta Line Deviation</td><td>0.000 nm</td></tr> <tr> <td>D2 Alpha Line Deviation</td><td>-0.146 nm</td></tr> <tr> <td>D2 Beta Line Deviation</td><td>-0.069 nm</td></tr> <tr> <td>Absolute D2 Alpha Line Deviation</td><td>-0.146 nm</td></tr> <tr> <td>Absolute D2 Beta Line Deviation</td><td>-0.069 nm</td></tr> <tr> <td>D2 Alpha Line</td><td>655.954 nm</td></tr> <tr> <td>D2 Beta Line</td><td>485.931 nm</td></tr> <tr> <td>Calibrate Detector with Wavelength Verificati</td><td>Yes</td></tr> </tbody> </table>	Name	Value	Cell Product Number	G4212-60011	Cell Name	Max-Light Test Cell	Cell Type	10 mm/0 µl	Lamp Type	Automatic Mode	Previous Absolute D2 Alpha Line Deviation	0.000 nm	Previous Absolute D2 Beta Line Deviation	0.000 nm	D2 Alpha Line Deviation	-0.146 nm	D2 Beta Line Deviation	-0.069 nm	Absolute D2 Alpha Line Deviation	-0.146 nm	Absolute D2 Beta Line Deviation	-0.069 nm	D2 Alpha Line	655.954 nm	D2 Beta Line	485.931 nm	Calibrate Detector with Wavelength Verificati	Yes
Name	Value																												
Cell Product Number	G4212-60011																												
Cell Name	Max-Light Test Cell																												
Cell Type	10 mm/0 µl																												
Lamp Type	Automatic Mode																												
Previous Absolute D2 Alpha Line Deviation	0.000 nm																												
Previous Absolute D2 Beta Line Deviation	0.000 nm																												
D2 Alpha Line Deviation	-0.146 nm																												
D2 Beta Line Deviation	-0.069 nm																												
Absolute D2 Alpha Line Deviation	-0.146 nm																												
Absolute D2 Beta Line Deviation	-0.069 nm																												
D2 Alpha Line	655.954 nm																												
D2 Beta Line	485.931 nm																												
Calibrate Detector with Wavelength Verificati	Yes																												

Figure 70 Wavelength Calibration – Results

Wavelength Recalibration Fails

Probable cause

- 1 Absorbing solvent or air bubble in Max-Light Cartridge Cell.
- 2 Dirty or contaminated Max-Light Cartridge Cell.
- 3 Old UV-lamp.
- 4 Dirty or contaminated optical components.

Suggested actions

- Repeat calibration with Max-Light Cartridge Test Cell and compare results.
- Ensure the Max-Light Cartridge Cell is filled with water.
 - Recalibrate.
- Exchange the UV-lamp.
- Run the Cell Test. If the test fails, flush the flow cell. See also [“Cleaning of Max-Light Cartridge Cell”](#) on page 218.

NOTE

If the test fails with Max-Light Cartridge Test Cell and new UV-lamp, the optical unit must be replaced.

D/A Converter (DAC) Test

The detector provides analog output of chromatographic signals for use with integrators, chart recorders or data systems. The analog signal is converted from the digital format by the digital-analog-converter (DAC).

The DAC test is used to verify correct operation of the digital-analog-converter by applying a digital test signal to the DAC.

The DAC outputs an analog signal of approximately 50 mV (if the zero offset of the analog output is set to the default value of 5 %) which can be plotted on an integrator. A continuous square wave with an amplitude of 10 μ V and a frequency of approximately 1 cycle/24 seconds is applied to the signal.

The amplitude of the square wave and the peak-to-peak noise are used to evaluate the DAC test.

When If the analog detector signal is noisy or missing.

Preparations Lamp must be on for at least 10 minutes. Connect integrator, chart recorder or data system to the detector analog output.

Running the test with Agilent LabAdvisor

- 1 Run the **D/A Converter (DAC) Test** (for further information see Online-Help of user interface).




Figure 71 D/A Converter (DAC) Test – Results




Figure 72 D/A Converter (DAC) Test – Example of Integrator Plot

Running the Test with Instant Pilot

The test can be started via the command line.

- 1 To start the test

TEST: DAC 1

Reply: **RA 00000 TEST:DAC 1**

- 2 To stop the test

TEST:DAC 0

Reply: **RA 00000 TEST:DAC 0**

Test Failed

D/A Converter (DAC) Test Evaluation

The noise on the step should be less than 3 μ V.

Probable cause

- 1 Bad cable or grounding problem between detector and external device.
- 2 Defective detector main board.

Suggested actions

- Check or replace the cable.
- Please contact your Agilent service representative.

Dark Current Test

The dark-current test measures the leakage current from each diode. The test is used to check for leaking diodes which may cause non-linearity at specific wavelengths. During the test, the slit assembly moves to the dark position, cutting off all light falling onto the diode array. Next, the leakage current from each diode is measured, and displayed graphically. The leakage current (represented in counts) for each diode should fall within the limits.

When

In case of problem.

- 1 Run the **Dark Current Test** with the recommended user interface (for further information see Online-Help of user interface).




Figure 73 Dark Current Test – Results




Figure 74 Dark Current Test – Signals

Test Failed

Dark-Current Test Evaluation

Probable cause

- 1 Defective slit assembly (stray light).
- 2 Defective detector main board.
- 3 Defective PDA/optical unit.

Suggested actions

- Run the “Self-Test” on page 178.
- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

9 Test Functions and Calibration

Dark Current Test



10 Maintenance

Warnings and Cautions	204
Introduction to Maintenance	206
Overview of Maintenance	207
Cleaning the Module	208
Replacing the Deuterium Lamp	209
Replacing the Max-Light Cartridge Cell	213
Cleaning of Max-Light Cartridge Cell	218
Storage of Max-Light Cartridge Cell	219
Drying the Leak Sensor	220
Replacing Leak Handling System Parts	221
Replacing Module Firmware	223
Information from Module's Assemblies	224

This chapter describes the maintenance of the module.



Warnings and Cautions

WARNING

Toxic, flammable and hazardous solvents, samples and reagents

The handling of solvents, samples and reagents can hold health and safety risks.

- When working with these substances observe appropriate safety procedures (for example by wearing goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet supplied by the vendor and follow good laboratory practice.
 - The amount of substances should be reduced to the minimal volume required for the analysis.
 - Do not operate the instrument in an explosive atmosphere.
-

WARNING

Eye damage by detector light



Eye damage may result from directly viewing the UV-light produced by the lamp of the optical system used in this product.

- Always turn the lamp of the optical system off before removing it.
-

WARNING

Electrical shock

Repair work at the module can lead to personal injuries, e.g. shock hazard, when the cover is opened.

- Do not remove the metal top cover of the module. No serviceable parts inside.
 - Only certified persons are authorized to carry out repairs inside the module.
-

WARNING

Personal injury or damage to the product

Agilent is not responsible for any damages caused, in whole or in part, by improper use of the products, unauthorized alterations, adjustments or modifications to the products, failure to comply with procedures in Agilent product user guides, or use of the products in violation of applicable laws, rules or regulations.

- Use your Agilent products only in the manner described in the Agilent product user guides.
-

CAUTION

Safety standards for external equipment

- If you connect external equipment to the instrument, make sure that you only use accessory units tested and approved according to the safety standards appropriate for the type of external equipment.
-

Introduction to Maintenance

The module is designed for easy maintenance. Maintenance can be done from the front with module in place in the system stack.

NOTE

There are no serviceable parts inside.

Do not open the module.

Overview of Maintenance

The following pages describe maintenance (simple repairs) of the detector that can be carried out without opening the main cover.

Table 25 Overview of Maintenance

Procedure	Typical Frequency	Notes
Cleaning of module	If required	
Deuterium lamp exchange	If noise and/or drift exceeds your application limits or lamp does not ignite.	A wavelength calibration test and an intensity test should be performed after replacement.
Flow cell exchange	If leaking or if intensity drops due to contaminated flow cell.	A wavelength calibration test should be performed after replacement.
Leak sensor drying	If leak has occurred.	Check for leaks.
Leak handling System replacement	If broken or corroded.	Check for leaks.

Cleaning the Module

The module case should be kept clean. Cleaning should be done with a soft cloth slightly dampened with water or a solution of water and mild detergent. Do not use an excessively damp cloth as liquid may drip into the module.

WARNING

Liquid dripping into the electronic compartment of your module.

Liquid in the module electronics can cause shock hazard and damage the module.

- Do not use an excessively damp cloth during cleaning.
 - Drain all solvent lines before opening any fittings.
-

Replacing the Deuterium Lamp

When If noise or drift exceeds application limits or lamp does not ignite.

Tools required Screwdriver POZI 1 PT3

Parts required	#	p/n	Description
	1	5190-0917	Long-life Deuterium lamp (8-pin) with RFID tag

Preparations Turn the lamp off.


WARNING

Injury by touching hot lamp


If the detector has been in use, the lamp may be hot.

→ If so, wait for lamp to cool down.

1 Remove the front cover.




2 Locate the lamp area.




10 Maintenance

Replacing the Deuterium Lamp


- 3** Unscrew the 3 screws of the lamp housing cover and remove the cover.




- 4** Locate the two screws that fix the lamp and unscrew.



- 5** Disconnect the lamp connector and remove the lamp.




- 6** Place the lamp on a clean place.




NOTE

Do not touch the glass bulb with your fingers. It may reduce the light output.

- 7** Insert the lamp and reconnect the lamp connector.



- 8** Locate the two screws and fix the lamp.




- 9** Fit the lamp wires in the lamp house cover so that they are not scratched by the cover.


10 Maintenance

Replacing the Deuterium Lamp

10 Replace the lamp housing cover and fix the 3 screws.



11 Close the front cover.



12 Perform a Wavelength Re-calibration after lamp warm-up.

Replacing the Max-Light Cartridge Cell

When If leaking or if intensity drops due to contaminated flow cell.

Tools required Hexagonal wrench


Parts required	#	p/n	Description
	1	G4212-60008	Max-Light Cartridge Flow Cell (10 mm, V(σ) 1.0 μ l)
	1	G4212-60007	Max-Light Cartridge Flow Cell (60 mm, V(σ) 4.0 μ L)
	1	G4212-60011	Max-Light Cartridge Test Cell

Preparations Turn the pump off.


NOTE

The flow cell is shipped with a filling of isopropanol. This is to avoid breakage due to subambient conditions. In case the flow cell is not used for some time (stored), then flush the flow cell with iso-propanol.

1 Remove the front cover.




2 Locate the cell area.




10 Maintenance

Replacing the Max-Light Cartridge Cell


- 3** Disconnect the inlet capillary to CELL-IN (left) and the waste tubing to CELL-OUT (right).




- 4** Unlock the cell cartridge holder by pulling the lever to the front.



- 5** The lever should be in the final down position.




- 6** Pull the cell cartridge holder completely out towards the front.




Replacing the Max-Light Cartridge Cell

- 7 Remove the cell from the cartridge holder.



- 8 Replace the black hoods [A] to the cell interfaces (in/out) and insert plugs [B] for save storage.


**NOTE**

The hoods and the plugs should be always in place to protect the flow cell.

For longer storage, the flow cell should be flushed and filled with isopropanol to prevent the grow of algae.

Store it in the plastic case provided with the Max-Light Cartridge Flow Cell.


- 9 Remove the black hoods from the cell interfaces (in/out) and insert the cell into the cell cartridge holder.




10 Maintenance

Replacing the Max-Light Cartridge Cell


10 Slide the cell cartridge holder completely into the module.




11 Lift the two levers into the upper final position to fix the cell.




12 Remove the plugs from the CELL-IN and CELL-OUT (keep them safe).



13 Connect the inlet capillary to CELL-IN (left) and the waste tubing to CELL-OUT (right).



14 Close the front cover.



Cleaning of Max-Light Cartridge Cell

When Low counts on Intensity Test or Cell Test (failed tests)

Tools required

- Alcohol (Iso-propanol or Ethanol)
- Lens tissue or Q-tips®
- Cell cleaning fluid, 1 L (p/n 5062-8529)

- 1 Flush the flow cell with the alcohol for some time.
- 2 Remove the cell from the cartridge holder (see [“Replacing the Max-Light Cartridge Cell”](#) on page 213).
- 3 Carefully clean the light inlet and outlet of the cell using lens tissue or Q-tips® with alcohol.

NOTE

If Q-tips® are used, ensure that no cotton fluff remains at the inlet or outlet.

NOTE

Do not touch the light inlet and outlet of the cell with your fingers. This will add a layer of contamination on the window and reduce the light throughput.

- 4 Flush the flow cell with water and repeat the Intensity Test and or Cell Test.
- 5 If tests fail again, the flow cell might be replaced if the chromatographic performance cannot be accepted.

NOTE

If the cleaning with the alcohol did not improve, you may use Cell cleaning fluid, 1 L (p/n 5062-8529).

Storage of Max-Light Cartridge Cell

- 1 Flush the Max-Light Cartridge Flow Cell with iso-propanol or methanol and insert the plugs into the cell inlet and outlet (see [“Replacing the Max-Light Cartridge Cell”](#) on page 213).
- 2 Remove the Max-Light Cartridge Cell from the cartridge holder of the detector.
- 3 Replace the black hoods, that secure the cell light inlet and outlet.
- 4 Store the Max-Light Cartridge Cell in plastic case provided with the Max-Light Cartridge Flow Cell.


Drying the Leak Sensor

When If leak has occurred.


Tools required Tissue

Preparations Turn the pump off.

1 Remove the front cover.




2 Locate the leak sensor area.



3 Dry the leak sensor and the area around. Check loose fittings at the flow cell. Note that the leak sensor does not touch the panel (clearance of about 1 mm).

4 Close the front cover.



Replacing Leak Handling System Parts


When If the parts are corroded or broken.

Tools required Tissue


Parts required	#	p/n	Description
	1	5061-3388	Leak funnel
	1	5041-8389	Leak funnel holder
	1	5062-2463	Tubing Flex 5 m
	1	G4212-40027	Leak downpipe

Preparations Turn the pump off.

1 Remove the front cover.




2 Locate the leak interface area.




10 Maintenance

Replacing Leak Handling System Parts


- 3** Pull the leak funnel out of the leak funnel holder and slide the Leak downpipe up for removal.



- 4** Insert the leak interface system parts. Ensure the tubing is fitted correctly in the bottom.



- 5** Close the front cover.



Replacing Module Firmware

When

The installation of newer firmware might be necessary

- if a newer version solves problems of older versions or
- to keep all systems on the same (validated) revision.

The installation of older firmware might be necessary

- to keep all systems on the same (validated) revision or
- if a new module with newer firmware is added to a system or
- if third part control software requires a special version.

Tools required

- LAN/RS-232 Firmware Update Tool or
- Agilent Lab Advisor software
- Instant Pilot G4208A (only if supported by module)

Parts required

#	Description
1	Firmware, tools and documentation from Agilent web site

Preparations

Read update documentation provided with the Firmware Update Tool.

To upgrade/downgrade the module’s firmware carry out the following steps:

- 1 Download the required module firmware, the latest LAN/RS-232 FW Update Tool and the documentation from the Agilent web.
 - http://www.chem.agilent.com/scripts/cag_firmware.asp.
- 2 To load the firmware into the module follow the instructions in the documentation.

Table 26 Module Specific Information (G4212A/B)

	G4212A - 1290 DAD	G4212B - 1260 DAD
Initial firmware (main and resident)	B.06.23	B.06.30
Compatibility with 1100/1200/1260/1290 series modules	When using the G4212A in a system, all other modules must have firmware revision A.06.1x or B.06.1x and above (main and resident). Otherwise the communication will not work.	When using the G4212B in a system, all other modules must have firmware revision A.06.3x or B.06.3x and above (main and resident). Otherwise the communication will not work.
Conversion to / emulation	N/A	N/A

Information from Module's Assemblies

Lamp and Flow Cell RFID Tag

The detector is equipped with a UV lamp and flow cell identification system using RFID (radio frequency identification) tags attached to the assemblies and RFID tag readers at the optical unit. The table below lists all parameters stored in the RFID tag.

Table 27 RFID Tag Data

Lamp information	Flow cell information
• product number	• product number
• serial number	• serial number
• production date	• production date
• accumulated UV on time (in hours)	• nominal path length of the cell (in mm)
• actual UV lamp on time (in hours)	• cell volume (σ) in μ l
• number of ignitions	• maximum pressure (in bar)
• date of last intensity test	• date of last cell test

NOTE

The pressure value is always displayed in bar, even if the user interface uses other units, e.g. PSI.

Serial Number and Firmware Revision

The user interface provides module specific information that is stored in the main board. These are for example the serial number, firmware revision.



11

Parts and Materials for Maintenance

Overview of Maintenance Parts [226](#)

Accessory Kits [228](#)

This chapter provides information on parts for maintenance.




Overview of Maintenance Parts

Item	p/n	Description
1	5190-0917	Long-life Deuterium lamp (8-pin) with RFID tag
2	G4212-60008	Max-Light Cartridge Flow Cell (10 mm, V(σ) 1.0 μ l)
3	G4212-60007	Max-Light Cartridge Flow Cell (60 mm, V(σ) 4.0 μ L)
4	G4212-60011	Max-Light Cartridge Test Cell
5	5067-4660	Inlet Capillary SST 0.12 mm I.D., 220 mm long
6	5062-2462	Teflon Tubing flexible i.d. 0.8 mm, o.d. 1.6 mm, 2 m, re-order 5 m (flow cell to waste)
7	5067-4691	Front Cover
8	5041-8388	Leak funnel
9	5041-8389	Leak funnel holder
10	5063-6527	Tubing assembly, i.d. 6 mm, o.d. 9 mm, 1.2 m (to waste)
11	G4212-40027	Leak downpipe
12	G4208A	Instant Pilot

NOTE

Instant Pilot G4208A (requires firmware B.02.11 or above).

For cables refer to [“Cable Overview”](#) on page 230.



Accessory Kits

Accessory kit (p/n G4212-68755) contains some specific accessories and tools needed for the installation of the detector.

p/n	Description
5062-2462	Teflon Tubing flexible i.d. 0.8 mm, o.d. 1.6 mm, 2 m, re-order 5 m (flow cell to waste)
5063-6527	Tubing assembly, i.d. 6 mm, o.d. 9 mm, 1.2 m (to waste)
5042-9954	Tubing clip (2x), re-order 4/pk
0100-1516	Fitting male PEEK, 2/pk
5067-4660	Inlet Capillary SST 0.12 mm I.D., 220 mm long
5181-1516	CAN cable, Agilent module to module, 0.5 m



12 Identifying Cables

Cable Overview	230
Analog Cables	232
Remote Cables	234
BCD Cables	237
CAN/LAN Cables	239
RS-232 Cable Kit	240

This chapter provides information on cables used with the Agilent 1260 Infinity/1290 Infinity LC modules.



Cable Overview

NOTE

Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.

Analog cables

p/n	Description
35900-60750	Agilent module to 3394/6 integrators
35900-60750	Agilent 35900A A/D converter
01046-60105	Analog cable (BNC to general purpose, spade lugs)

Remote cables

p/n	Description
03394-60600	Agilent module to 3396A Series I integrators 3396 Series II / 3395A integrator, see details in section “Remote Cables” on page 234
03396-61010	Agilent module to 3396 Series III / 3395B integrators
5061-3378	Agilent module to Agilent 35900 A/D converters (or HP 1050/1046A/1049A)
01046-60201	Agilent module to general purpose

BCD cables

p/n	Description
03396-60560	Agilent module to 3396 integrators
G1351-81600	Agilent module to general purpose

CAN cables

p/n	Description
5181-1516	CAN cable, Agilent module to module, 0.5 m
5181-1519	CAN cable, Agilent module to module, 1 m


LAN cables

p/n	Description
5023-0203	Cross-over network cable, shielded, 3 m (for point to point connection)
5023-0202	Twisted pair network cable, shielded, 7 m (for point to point connection)

RS-232 cables


p/n	Description
G1530-60600	RS-232 cable, 2 m
RS232-61600	RS-232 cable, 2.5 m Instrument to PC, 9-to-9 pin (female). This cable has special pin-out, and is not compatible with connecting printers and plotters. It's also called "Null Modem Cable" with full handshaking where the wiring is made between pins 1-1, 2-3, 3-2, 4-6, 5-5, 6-4, 7-8, 8-7, 9-9.
5181-1561	RS-232 cable, 8 m

Analog Cables




One end of these cables provides a BNC connector to be connected to Agilent modules. The other end depends on the instrument to which connection is being made.


Agilent Module to 3394/6 Integrators

p/n 35900-60750	Pin 3394/6	Pin Agilent module	Signal Name
	1		Not connected
	2	Shield	Analog -
	3	Center	Analog +


Agilent Module to BNC Connector

p/n 8120-1840	Pin BNC	Pin Agilent module	Signal Name
	Shield	Shield	Analog -
	Center	Center	Analog +

Agilent Module to General Purpose

p/n 01046-60105	Pin 3394/6	Pin Agilent module	Signal Name
	1		Not connected
	2	Black	Analog -
	3	Red	Analog +

Remote Cables



One end of these cables provides a Agilent Technologies APG (Analytical Products Group) remote connector to be connected to Agilent modules. The other end depends on the instrument to be connected to.


Agilent Module to 3396A Integrators

p/n 03394-60600	Pin 3394	Pin Agilent module	Signal Name	Active (TTL)
	9	1 - White	Digital ground	
	NC	2 - Brown	Prepare run	Low
	3	3 - Gray	Start	Low
	NC	4 - Blue	Shut down	Low
	NC	5 - Pink	Not connected	
	NC	6 - Yellow	Power on	High
	5,14	7 - Red	Ready	High
	1	8 - Green	Stop	Low
	NC	9 - Black	Start request	Low
	13, 15		Not connected	


Agilent Module to 3396 Series II / 3395A Integrators

Use the cable Agilent module to 3396A Series I integrators (p/n 03394-60600) and cut pin #5 on the integrator side. Otherwise the integrator prints START; not ready.


Agilent Module to 3396 Series III / 3395B Integrators

p/n 03396-61010	Pin 33XX	Pin Agilent module	Signal Name	Active (TTL)
	9	1 - White	Digital ground	
	NC	2 - Brown	Prepare run	Low
	3	3 - Gray	Start	Low
	NC	4 - Blue	Shut down	Low
	NC	5 - Pink	Not connected	
	NC	6 - Yellow	Power on	High
	14	7 - Red	Ready	High
	4	8 - Green	Stop	Low
	NC	9 - Black	Start request	Low
	13, 15		Not connected	

Agilent Module to Agilent 35900 A/D Converters

p/n 5061-3378	Pin 35900 A/D	Pin Agilent module	Signal Name	Active (TTL)
	1 - White	1 - White	Digital ground	
	2 - Brown	2 - Brown	Prepare run	Low
	3 - Gray	3 - Gray	Start	Low
	4 - Blue	4 - Blue	Shut down	Low
	5 - Pink	5 - Pink	Not connected	
	6 - Yellow	6 - Yellow	Power on	High
	7 - Red	7 - Red	Ready	High
	8 - Green	8 - Green	Stop	Low
	9 - Black	9 - Black	Start request	Low

Agilent Module to General Purpose


p/n 01046-60201	Pin Universal	Pin Agilent module	Signal Name	Active (TTL)
		1 - White	Digital ground	
		2 - Brown	Prepare run	Low
		3 - Gray	Start	Low
		4 - Blue	Shut down	Low
		5 - Pink	Not connected	
		6 - Yellow	Power on	High
		7 - Red	Ready	High
		8 - Green	Stop	Low
		9 - Black	Start request	Low

BCD Cables




One end of these cables provides a 15-pin BCD connector to be connected to the Agilent modules. The other end depends on the instrument to be connected to

Agilent Module to General Purpose


p/n G1351-81600	Wire Color	Pin Agilent module	Signal Name	BCD Digit
	Green	1	BCD 5	20
	Violet	2	BCD 7	80
	Blue	3	BCD 6	40
	Yellow	4	BCD 4	10
	Black	5	BCD 0	1
	Orange	6	BCD 3	8
	Red	7	BCD 2	4
	Brown	8	BCD 1	2
	Gray	9	Digital ground	Gray
	Gray/pink	10	BCD 11	800
	Red/blue	11	BCD 10	400
	White/green	12	BCD 9	200
	Brown/green	13	BCD 8	100
	not connected	14		
	not connected	15	+ 5 V	Low

12 Identifying Cables
BCD Cables

Agilent Module to 3396 Integrators

p/n 03396-60560	Pin 3396	Pin Agilent module	Signal Name	BCD Digit
	1	1	BCD 5	20
	2	2	BCD 7	80
	3	3	BCD 6	40
	4	4	BCD 4	10
	5	5	BCD0	1
	6	6	BCD 3	8
	7	7	BCD 2	4
	8	8	BCD 1	2
	9	9	Digital ground	
	NC	15	+ 5 V	Low

CAN/LAN Cables



Both ends of this cable provide a modular plug to be connected to Agilent modules CAN or LAN connectors.

CAN Cables

p/n	Description
5181-1516	CAN cable, Agilent module to module, 0.5 m
5181-1519	CAN cable, Agilent module to module, 1 m

LAN Cables

p/n	Description
5023-0203	Cross-over network cable, shielded, 3 m (for point to point connection)
5023-0202	Twisted pair network cable, shielded, 7 m (for point to point connection)

RS-232 Cable Kit

p/n	Description
G1530-60600	RS-232 cable, 2 m
RS232-61600	RS-232 cable, 2.5 m Instrument to PC, 9-to-9 pin (female). This cable has special pin-out, and is not compatible with connecting printers and plotters. It's also called "Null Modem Cable" with full handshaking where the wiring is made between pins 1-1, 2-3, 3-2, 4-6, 5-5, 6-4, 7-8, 8-7, 9-9.
5181-1561	RS-232 cable, 8 m



13 Appendix

Safety [242](#)

The Waste Electrical and Electronic Equipment Directive [246](#)

Radio Interference [247](#)

Sound Emission [248](#)

Flow Cell [249](#)

Agilent Technologies on Internet [250](#)






This chapter provides addition information on safety, legal and web.



Safety

Safety Symbols

Table 28 Safety Symbols

Symbol	Description
	The apparatus is marked with this symbol when the user should refer to the instruction manual in order to protect risk of harm to the operator and to protect the apparatus against damage.
	Indicates dangerous voltages.
	Indicates a protected ground terminal.
	Indicates eye damage may result from directly viewing the light produced by the deuterium lamp used in this product.
	The apparatus is marked with this symbol when hot surfaces are available and the user should not touch it when heated up.

WARNING

A WARNING

alerts you to situations that could cause physical injury or death.

- Do not proceed beyond a warning until you have fully understood and met the indicated conditions.

CAUTION

A CAUTION

alerts you to situations that could cause loss of data, or damage of equipment.

- Do not proceed beyond a caution until you have fully understood and met the indicated conditions.

General Safety Information

The following general safety precautions must be observed during all phases of operation, service, and repair of this instrument. Failure to comply with these precautions or with specific warnings elsewhere in this manual violates safety standards of design, manufacture, and intended use of the instrument. Agilent Technologies assumes no liability for the customer's failure to comply with these requirements.

WARNING

Ensure the proper usage of the equipment.

The protection provided by the equipment may be impaired.

→ The operator of this instrument is advised to use the equipment in a manner as specified in this manual.

Safety Standards

This is a Safety Class I instrument (provided with terminal for protective earthing) and has been manufactured and tested according to international safety standards.

Operation

Before applying power, comply with the installation section. Additionally the following must be observed.

Do not remove instrument covers when operating. Before the instrument is switched on, all protective earth terminals, extension cords, auto-transformers, and devices connected to it must be connected to a protective earth via a ground socket. Any interruption of the protective earth grounding will cause a potential shock hazard that could result in serious personal injury. Whenever it is likely that the protection has been impaired, the instrument must be made inoperative and be secured against any intended operation.

Make sure that only fuses with the required rated current and of the specified type (normal blow, time delay, and so on) are used for replacement. The use of repaired fuses and the short-circuiting of fuse holders must be avoided.

Some adjustments described in the manual, are made with power supplied to the instrument, and protective covers removed. Energy available at many points may, if contacted, result in personal injury.

Any adjustment, maintenance, and repair of the opened instrument under voltage should be avoided whenever possible. When inevitable, this has to be carried out by a skilled person who is aware of the hazard involved. Do not attempt internal service or adjustment unless another person, capable of rendering first aid and resuscitation, is present. Do not replace components with power cable connected.

Do not operate the instrument in the presence of flammable gases or fumes. Operation of any electrical instrument in such an environment constitutes a definite safety hazard.

Do not install substitute parts or make any unauthorized modification to the instrument.

Capacitors inside the instrument may still be charged, even though the instrument has been disconnected from its source of supply. Dangerous voltages, capable of causing serious personal injury, are present in this instrument. Use extreme caution when handling, testing and adjusting.

When working with solvents please observe appropriate safety procedures (e.g. goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet by the solvent vendor, especially when toxic or hazardous solvents are used.

The Waste Electrical and Electronic Equipment Directive

Abstract

The Waste Electrical and Electronic Equipment (WEEE) Directive (2002/96/EC), adopted by EU Commission on 13 February 2003, is introducing producer responsibility on all electric and electronic appliances starting with 13 August 2005.

NOTE

This product complies with the WEEE Directive (2002/96/EC) marking requirements. The affixed label indicates that you must not discard this electrical/electronic product in domestic household waste.

Product Category:

With reference to the equipment types in the WEEE Directive Annex I, this product is classed as a Monitoring and Control Instrumentation product.



NOTE

Do not dispose off in domestic household waste

To return unwanted products, contact your local Agilent office, or see www.agilent.com for more information.

Radio Interference

Cables supplied by Agilent Technologies are screened to provide optimized protection against radio interference. All cables are in compliance with safety or EMC regulations.

Test and Measurement

If test and measurement equipment is operated with unscreened cables, or used for measurements on open set-ups, the user has to assure that under operating conditions the radio interference limits are still met within the premises.

Sound Emission

Manufacturer's Declaration

This statement is provided to comply with the requirements of the German Sound Emission Directive of 18 January 1991.

This product has a sound pressure emission (at the operator position) < 70 dB.

- Sound Pressure $L_p < 70$ dB (A)
- At Operator Position
- Normal Operation
- According to ISO 7779:1988/EN 27779/1991 (Type Test)

Flow Cell


To protect optimal functionality of your flow-cell:

- Avoid the use of alkaline solutions (pH > 9.5) which can attack quartz and thus impair the optical properties of the flow cell.

Use of Solvents

Observe the following recommendations on the use of solvents.

- Brown glass ware can avoid growth of algae.
- Avoid the use of the following steel-corrosive solvents:
 - Solutions of alkali halides and their respective acids (for example, lithium iodide, potassium chloride, and so on),
 - High concentrations of inorganic acids like sulfuric acid and nitric acid, especially at higher temperatures (if your chromatography method allows, replace by phosphoric acid or phosphate buffer which are less corrosive against stainless steel),
 - Halogenated solvents or mixtures which form radicals and/or acids, for example:



This reaction, in which stainless steel probably acts as a catalyst, occurs quickly with dried chloroform if the drying process removes the stabilizing alcohol,

- Chromatographic grade ethers, which can contain peroxides (for example, THF, dioxane, di-isopropylether) such ethers should be filtered through dry aluminium oxide which adsorbs the peroxides,
- Solvents containing strong complexing agents (e.g. EDTA),
- Mixtures of carbon tetrachloride with 2-propanol or THF.

Agilent Technologies on Internet

For the latest information on products and services visit our worldwide web site on the Internet at:

<http://www.agilent.com>

Select Products/Chemical Analysis

It will provide also the latest firmware of the modules for download.

Index

8

- 8-bit configuration switch
 - on-board LAN 28
 - without On-Board LAN 30

A

- accessory kit parts 228
- accessory kit 47
- advanced method parameter settings
 - using 110
- Agilent Diagnostic software 156
- Agilent Lab Advisor software 156
- Agilent Lab Advisor 156
- Agilent
 - ChemStation Setup 90
 - on internet 250
- algae 249
- ambient operating temperature 40
- ambient non-operating temperature 40
- analog signal output 42
- analog signal 25
- analog
 - cable 232
- apg remote 26
- array
 - diode 16
- astm drift 188
- ASTM
 - environmental conditions 39
- automatic configuration with Bootp 74

B

- bandwidth 145

- BCD
 - cable 237
- Beer-Lambert (law) 127
- bench space 38
- BootP service
 - installation 76
 - restart 83
 - settings 82
 - stop 82
- Bootp
 - & Store 70
 - automatic configuration 74
 - initialization modes 69
 - using default 71
 - using stored 71

C

- cable
 - analog 232
 - BCD 237
 - CAN 239
 - connecting APG remote 54
 - connecting CAN 54
 - connecting GPIB 54
 - connecting LAN 54
 - connecting the ChemStation 54
 - connecting the power 54
 - LAN 239
 - remote 234
 - RS-232 240
- cables
 - analog 230
 - BCD 230
 - CAN 231
 - LAN 231
 - overview 230
 - remote 230
 - RS-232 231
- CAN
 - cable 239
- cartridge
 - cleaning 218
 - max-light 218
 - storage 219
- cautions and warnings 204
- cell test 183
- ChemStation
 - setup 93
- cleaning
 - flow cell 218
- Communication settings
 - RS-232C 31
- compensation sensor open 162
- compensation sensor short 163
- condensation 39, 46
- configuration switch 68
- configuration
 - one stack 49, 49, 51
 - stack 48
 - two stack front 55
 - two stack rear 56
 - two stack 53, 55
- control and data evaluation 42
- control settings
 - using 106
- correction factors for flow cells 128

D

- D/A converter 198

Index

- DAC
 - Agilent LabAdvisor 198
 - Instant Pilot 199
- dark current 200
- data collection rate 147
- data evaluation and control 42
- defect on arrival 46
- delay volume 142
- delivery checklist 47
- detection
 - compound classes 140
- detector warm up 149
- detector
 - achieving higher sensitivity 143
 - installation 57
 - preparation 100
 - setup with ChemStation 101
- Diagnostic software 156
- dimensions 40
- diode
 - width 41
- drift (ASTM) and noise 41
- E**
- electrical connections
 - descriptions of 19
- electronic waste 246
- EMF
 - early maintenance feedback 17
- environment 39
- error messages
 - fan failed 163
 - uv heater current 168
 - uv ignition failed 167
- error messages
 - compensation sensor open 162
 - compensation sensor short 163
 - cover violation 165
 - diode current leakage 166
 - heater power at limit 171
 - heater failed 170
 - ignition without cover 164, 164
 - illegal value from air inlet sensor 170
 - illegal value from heater sensor 169
 - leak sensor open 162
 - leak sensor short 161
 - leak 164
 - remote timeout 160
 - shut-down 160
 - synchronization lost 161
 - timeout 159
 - uv lamp current 166
 - uv lamp voltage 167
 - wavelength calibration failed 168
 - wavelength recalibration lost 169
- extra-column volume
 - description 142
- F**
- fan failed 163
- features
 - safety and maintenance 42
- firmware
 - updates 223
 - upgrade/downgrade 223
- flow cell
 - max -light cartridge 13
 - solvent information 249
- flow cell
 - correction factors 128
 - Max-Light cartridge flow cell 143
 - Max-Light high sensitivity cell 143
 - specifications 41
- flow connections 60, 60
- frequency range 40
- G**
- general error messages 159
- general method settings
 - using 108
- GLP features 42
- grating 16
- GUI
 - detector 103
- H**
- humidity 40
- I**
- initialization mode selection 69
- installation
 - accessory kit 47
 - bench space 38
 - delivery checklist 47
 - detector 57
 - environment 39
 - flow connections 60, 60
 - of the detector 57
 - power considerations 36
- Instant Pilot
 - using 119
- instrument configuration
 - using 117
- instrument curves
 - using 116
- instrument layout 18
- intensity test 180
- interfaces 22
- internet 250
- L**
- lamp
 - uv 12
- LAN configuration
 - ChemStation 93
 - PC setup 90
- LAN

Index

- automatic configuration with
 - Bootp 74
 - Bootp & Store 70
 - Bootp 69
 - cable 239
 - configuration switch 68
 - first steps 66
 - initialization mode selection 69
 - link configuration selection 73
 - manual configuration with telnet 85
 - manual configuration 84
 - PC and Agilent ChemStation Setup 90
 - TCP/IP parameter configuration 67
 - using default 71
 - using stored 71
 - what you have to do first 66
 - leak sensor open 162
 - leak sensor short 161
 - leak 164
 - line frequency 40
 - line voltage 40
 - linear range 41
 - linearity
 - specifications 43
 - link configuration selection 73
 - local configuration 90
- ## M
- MAC address
 - determine 79
 - MAC
 - address 66
 - maintenance
 - definition of 206
 - drying the leak sensor 220
 - feedback 17
 - lamp and flow cell RFID tag 224
 - overview 207
 - replacing firmware 223
 - replacing leak handling system 221
 - replacing the deuterium lamp 209
 - replacing the flow cell 213
 - serial number and firmware revision 224
 - manual configuration
 - of LAN 84
 - max-light cartridge
 - flow cell 13
 - message
 - cover violation 165
 - diode current leakage 166
 - heater failed 170
 - heater power at limit 171
 - ignition without cover 164, 164
 - illegal value from air inlet senso 170
 - illegal value from heater sensor 169
 - remote timeout 160
 - uv heater current 168
 - uv ignition failed 167
 - uv lamp current 166
 - uv lamp voltage 167
 - wavelength calibration failed 168
 - method parameter settings
 - using 107
- ## N
- negative absorbance 137
 - noise and drift (ASTM) 41
 - Noise and Linearity
 - specifications 43
 - noise 188
 - non-operating altitude 40
 - non-operating temperature 40
- ## O
- operating Altitude 40
 - operating temperature 40
 - optical 11
- ## Optimization
- optimization
 - detector performance 124
 - detector sensitivity 143
 - for sensitivity, selectivity, linearity, dispersion 127
 - how to get the best performance 124
 - margins for negative absorbance 137
 - of selectivity 138
 - overview 125
 - peak width 129
 - performance 123
 - sample and reference wavelength 131
 - slit width 134
 - spectra acquisition 136
 - stack configuration 48
 - wavelength and bandwidth 144
 - optimizing
 - 1260 System 142
 - flow cell 126
 - other advanced method parameter using 113
 - overview for maintenance 226
- ## P
- packaging
 - damaged 46
 - parts identification
 - accessory kit 228
 - PC and Agilent ChemStation Setup 90
 - PC setup
 - local configuration 90
 - peak width (response time) 129
 - peak width 147
 - performance
 - optimization 124
 - specifications 41
 - photometric accuracy 128

Index

physical specifications 40
power supply indicator 153
power considerations 36
power consumption 40
power cords 37
programmable slit width 41

Q

quick noise test 185

R

radio interference 247
recalibration
 initial 64
recyclable material 42
remote
 cable 234
repairs
 cautions and warnings 204
 replacing firmware 223
response time (peak width) 129
response time versus time constant 43
response time 147
RS-232C
 cable 240
 communication settings 31

S

safety class I 243
safety
 general information 243
 standards 40
 symbols 242
sample and reference wavelength 131
selectivity optimization 138
self test 178
serial number
 information 21, 21

setup
 ChemStation 93
 detector 101
shut-down 160
signal wavelength 145
site requirements
 power cords 37
slit test 191
slit width 41, 134
slit
 fix 15
 programmable 14
solvents 249
sound emission 248
special interfaces 27
special settings
 boot-resident 33
 forced cold start 34
specification
 physical 40
specifications
 analog signal output 42
 communications 42
 control and data evaluation 42
 diode width 41
 flow cell 41
 GLP features 42
 linear range 41
 noise and drift (ASTM) 41
 noise and linearity 43
 programmable slit width 41
 safety and maintenance 42
 signal data rate 41
 spectra data rate 41
 wavelength accuracy 41
 wavelength bunching 41
 wavelength range 41
spectra
 acquisition 136
spectrum settings

 using 111
stack configuration 54
 rear view 54
suppression
 quantifying 138
synchronization lost 161
system setup and installation
 optimizing stack configuration 48
system 11

T

TCP/IP parameter configuration 67
telnet
 configuration 85
temperature sensor 164
test function
 astm drift 188
 D/A converter 198
test functions 152
test cell
 usage 175
test function
 cell test 183
 conditions 176
 DAC 198
 dark current 200
 failing 177
 intensity test 180
 introduction 174
 max-light cartridge 175
 noise 188
 quick noise test 185
 self test 178
 slit 191
 test cell cartridge 175
 wavelength calibration 195
 wavelength verification 193
time table settings
 using 114
time constant versus response time 43

Index

timeout 159

troubleshooting

- error messages 152, 158
- status indicators 152, 153

U

unpacking 46

user interfaces 155

using

- advanced method parameter settings 110
- control settings 106
- detector 100
- general method settings 108
- GUI 103
- Instant Pilot 119
- instrument configuration 117
- instrument curves 116
- method parameter settings 107
- other advanced method parameter 113
- spectrum settings 111
- time table settings 114

wavelength

- accuracy 41
- bunching 41
- range 41

WEEE directive 246

weight 40

V

voltage range 40

W

warm up of detector 149

warnings and cautions 204

waste

- electrical and electronic equipment 246

wavelength and bandwidth

- optimization 144

wavelength recalibration lost 169

wavelength calibration 195

wavelength verification

- test 193

In This Book

This manual contains technical reference information about the Agilent 1290 Infinity Diode Array Detector (G4212A) and the Agilent 1260 Infinity Diode Array Detector (G4212B).

- introduction and specifications,
- installation,
- using and optimizing,
- troubleshooting and diagnose,
- maintenance,
- parts identification,
- safety and related information.

© Agilent Technologies 2010

Printed in Germany
06/2010



G4212-90010



Agilent Technologies